



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112629

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/05/2021	14:25
2	M1010-02		1	8	100.00	>212.0	>212.0	01/05/2021	14:45
3	M1010-02DUP		1	9	100.00	>212.0	>212.0	01/05/2021	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB112629

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP1521 WorkList ID : 145400 Department : Wet-Chemistry Date : 01-05-2021 13:59:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2021	Solid	M1010-02	Flash Point	Cool 4 deg C	PSEG03	M12	MT	01/05/2021	1010A

Date/Time 01/05/21 14:00
Received by: KS (Weichen)
Relinquished by: AP C Ext. 467

Date/Time 01/05/21 15:25
Received by: AP C Ext. 467
Relinquished by: KS (Weichen)