


LB112630

INSTRUMENT ID: CV4

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	49	0.000	49				
0.2	0.200	0.188	-0.012	752	0.0 %	752				
2.5	2.500	2.447	-0.053	9078	0.0 %	9078				
5.0	5.000	5.214	0.214	19278	0.0 %	19278				
7.5	7.500	7.321	-0.179	27045	0.0 %	27045				
10.0	10.000	10.033	0.033	37043	0.0 %	37043				

LB112630 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Jan 2021 15:22:53				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	49	-	0.00007470A	PPB	05 Jan 2021 15:24:14	S	Std
0.2 - 1	50.2	752	-	0.20007470A	PPB	05 Jan 2021 15:26:31	S	Std
2.5 - 1	42.5	9078	-	2.50007470A	PPB	05 Jan 2021 15:28:48	S	Std
5.0 - 1	55	19278	-	5.00007470A	PPB	05 Jan 2021 15:31:05	S	Std
7.5 - 1	57.5	27045	-	7.50007470A	PPB	05 Jan 2021 15:33:21	S	Std
10.0 - 1	51.6	37043	-	10.00007470A	PPB	05 Jan 2021 15:35:37	S	Std
ICV60 - 1	ICV60	15228	4.1151	-7470A	PPB	05 Jan 2021 15:38:15	U	SMPL
ICB60 - 1	ICB60	-37	-0.0258	-7470A	PPB	05 Jan 2021 15:40:30	U	SMPL
CCV72 - 1	CCV72	19634	5.3104	-7470A	PPB	05 Jan 2021 15:42:45	U	SMPL
CCB72 - 1	CCB72	-41	-0.0269	-7470A	PPB	05 Jan 2021 15:45:02	U	SMPL
CRA - 1	CRA	717	0.1787	-7470A	PPB	05 Jan 2021 15:47:17	U	SMPL
HighStd - 1	HighStd	36700	9.9399	-7470A	PPB	05 Jan 2021 15:49:32	U	SMPL
ChkStd - 1	ChkStd	24269	6.5677	-7470A	PPB	05 Jan 2021 15:51:48	U	SMPL
PB133921BL - 1	PBW	-78	-0.0369	-7470A	PPB	05 Jan 2021 15:54:04	U	SMPL
PB133921BS - 1	LCSW	13625	3.6803	-7470A	PPB	05 Jan 2021 15:56:20	U	SMPL
L5126-17 - 1	20071-OILY-DEBRIS-COMP	-22	-0.0217	-7470A	PPB	05 Jan 2021 15:58:36	U	SMPL
L5192-05 - 1	CL-02-122220	18	-0.0109	-7470A	PPB	05 Jan 2021 16:00:53	U	SMPL
L5192-05DUP - 1	CL-02-122220DUP	24	-0.0093	-7470A	PPB	05 Jan 2021 16:03:10	U	SMPL
L5192-05MS - 1	CL-02-122220MS	13677	3.6944	-7470A	PPB	05 Jan 2021 16:05:27	U	SMPL
L5192-05MSD - 1	CL-02-122220MSD	14521	3.9233	-7470A	PPB	05 Jan 2021 16:07:43	U	SMPL
CCV73 - 1	CCV73	19407	5.2488	-7470A	PPB	05 Jan 2021 16:12:50	U	SMPL
CCB73 - 1	CCB73	-37	-0.0258	-7470A	PPB	05 Jan 2021 16:15:05	U	SMPL
PB133921TB - 1	PB133895TB	18	-0.0109	-7470A	PPB	05 Jan 2021 16:17:20	U	SMPL
L5192-05LX5 - 1		12	-0.0125	-7470A	PPB	05 Jan 2021 16:19:36	U	SMPL
L5192-05A - 1		14735	3.9814	-7470A	PPB	05 Jan 2021 16:21:51	U	SMPL
CCV74 - 1	CCV74	19215	5.1967	-7470A	PPB	05 Jan 2021 16:24:07	U	SMPL
CCB74 - 1	CCB74	-38	-0.0261	-7470A	PPB	05 Jan 2021 16:26:23	U	SMPL