



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112634

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/05/2021	16:20
2	CAL2	1	Water	NA	NA	20.2	7.01	01/05/2021	16:21
3	CAL3	1	Water	NA	NA	20.2	10.03	01/05/2021	16:22
4	ICV	1	Water	NA	NA	20.2	6.98	01/05/2021	16:24
5	CCV1	1	Water	NA	NA	20.1	2.01	01/05/2021	16:25
6	M1010-02	1	Solid	20.02	20	20.2	5.12	01/05/2021	16:26
7	M1010-02DUP	1	Solid	20.03	20	20.3	5.13	01/05/2021	16:27
8	CCV2	1	Water	NA	NA	20.1	12.02	01/05/2021	16:30

27 112634

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1008 WorkList ID : 145407 Department : Wet-Chemistry Date : 01-05-2021 15:56:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/05/2021	Solid	M1010-02	Corrosivity	Cool 4 deg C	PSEG03	M12	MT	01/05/2021	9045D

Date/Time 01-05-2021 16:00
Received by: J.D. (wet Chem)
Relinquished by: KS (wetchem)

Date/Time 01-05-2021 17:10
Received by: KS (wetchem)
Relinquished by: J.D. (wet Chem)