



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112635

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/05/2021	16:20
2	CAL2	1	Water	NA	NA	20.2	7.01	01/05/2021	16:21
3	CAL3	1	Water	NA	NA	20.2	10.03	01/05/2021	16:22
4	ICV	1	Water	NA	NA	20.2	6.98	01/05/2021	16:24
5	CCV1	1	Water	NA	NA	20.1	2.01	01/05/2021	16:25
6	M1011-01	1	Water	NA	NA	13.8	11.70	01/05/2021	16:34
7	M1011-01DUP	1	Water	NA	NA	13.9	11.71	01/05/2021	16:35
8	CCV2	1	Water	NA	NA	20.2	12.02	01/05/2021	16:36

WORKLIST(Hardcopy Internal Chain)

112635

WorkList Name : CORROSIVITY M1011 V WorkList ID : 145410 Department : Wet-Chemistry Date : 01-05-2021 16:14:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/05/2021	Water	M1011-01	Corrosivity	Cool 4 deg C	PSEG03	M12	N48967	01/05/2021	9040C

Date/Time 01-05-2021 16:14:29
 Received by: JQ (wet Chem)
 Relinquished by: KS (wet Chem)

Date/Time 01-05-2021 17:25
 Received by: KS (wet Chem)
 Relinquished by: JQ (wet Chem)