



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112636

Slope : 97.6

pH Meter ID : WC PH METER-1

| Calibration Standards | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION | W2718 |
| buffer solution pH 7 yellow | W2627 |
| PH 10.01 BUFFER,COLOR CD 475ML | W2633 |
| buffer solution pH 7 yellow | W2719 |
| Buffer Solution, PH2 (500ml) | W2592 |
| Buffer Solution, PH12 (500ml) | W2616 |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1 | CAL1 | 1 | Water | NA | NA | 20.2 | 4.01 | 01/05/2021 | 16:20 |
| 2 | CAL2 | 1 | Water | NA | NA | 20.2 | 7.01 | 01/05/2021 | 16:21 |
| 3 | CAL3 | 1 | Water | NA | NA | 20.2 | 10.03 | 01/05/2021 | 16:22 |
| 4 | ICV | 1 | Water | NA | NA | 20.2 | 6.98 | 01/05/2021 | 16:24 |
| 5 | CCV1 | 1 | Water | NA | NA | 20.1 | 2.01 | 01/05/2021 | 16:25 |
| 6 | M1011-02 | 1 | Water | NA | NA | 13.8 | 11.70 | 01/05/2021 | 16:34 |
| 7 | M1011-02DUP | 1 | Water | NA | NA | 13.9 | 11.71 | 01/05/2021 | 16:35 |
| 8 | CCV2 | 1 | Water | NA | NA | 20.2 | 12.02 | 01/05/2021 | 16:36 |

28 112636

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1008 WorkList ID : 145406 Department : Wet-Chemistry Date : 01-05-2021 15:56:08

| Due Date | Matrix | Sample | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|------------------|-----------------|--------------|--------|
| 01/05/2021 | Water | M1011-02 | pH | Cool 4 deg C | PSEG03 | M12 | N48967 | 01/05/2021 | 9040C |

Date/Time 01.05.2021 16:15
 Received by: JS (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 01.05.2021 17:25
 Received by: KS (Wet Chem)
 Relinquished by: SP (Wet Chem)