


LB112641

SFAM01.1

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.1092e-004

C= -5.9802e-003

Rho= 0.9999967

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.00	0.000	0.014	0.014	95	0.000	95					-
0.05	0.050	0.047	-0.003	249	0.0 %	249					-6
0.20	0.200	0.201	0.001	982	0.0 %	982					1
2.50	2.500	2.490	-0.010	11832	0.0 %	11832					0
5.00	5.000	4.998	-0.002	23726	0.0 %	23726					0
7.50	7.500	7.488	-0.012	35528	0.0 %	35528					0
10.0	10.000	10.013	0.013	47499	0.0 %	47499					0

LB112641
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	95	-		0 SFAM01.1	PPB	1/6/2021 11:09	Std
	0.05 S01	249	-	0.05	SFAM01.1	PPB	1/6/2021 11:11	Std
	0.2 S02	982	-	0.2	SFAM01.1	PPB	1/6/2021 11:14	Std
	2.5 S03	11832	-	2.5	SFAM01.1	PPB	1/6/2021 11:16	Std
	5 S04	23726	-	5	SFAM01.1	PPB	1/6/2021 11:18	Std
	7.5 S05	35528	-	7.5	SFAM01.1	PPB	1/6/2021 11:21	Std
	10 S06	47499	-	10	SFAM01.1	PPB	1/6/2021 11:25	Std
ICV006	ICV006	19129	4.0287	-	SFAM01.1	PPB	1/6/2021 11:28	SMPL
ICB006	ICB006	37	0.0018	-	SFAM01.1	PPB	1/6/2021 11:30	SMPL
CCV075	CCV075	23879	5.0306	-	SFAM01.1	PPB	1/6/2021 11:33	SMPL
CCB075	CCB075	27	-0.0003	-	SFAM01.1	PPB	1/6/2021 11:35	SMPL
PB133944BL	PBW944	76	0.01	-	SFAM01.1	PPB	1/6/2021 11:37	SMPL
L5214-04	MDL-SOIL-QT1-2021-02	259	0.0486	-	SFAM01.1	PPB	1/6/2021 11:42	SMPL
CCV076	CCV076	23879	5.0306	-	SFAM01.1	PPB	1/6/2021 11:44	SMPL
CCB076	CCB076	48	0.0041	-	SFAM01.1	PPB	1/6/2021 11:47	SMPL