



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112642

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/06/2021	11:22
2	CAL2	1	Water	NA	NA	20.2	7.00	01/06/2021	11:23
3	CAL3	1	Water	NA	NA	20.2	10.03	01/06/2021	11:25
4	ICV	1	Water	NA	NA	20.2	6.98	01/06/2021	11:26
5	CCV1	1	Water	NA	NA	20.2	2.01	01/06/2021	11:30
6	M1018-01	1	Water	NA	NA	10.2	6.95	01/06/2021	11:32
7	M1018-01DUP	1	Water	NA	NA	10.3	6.96	01/06/2021	11:31
8	M1018-03	1	Water	NA	NA	10.1	7.02	01/06/2021	11:34
9	CCV2	1	Water	NA	NA	20.2	12.02	01/06/2021	11:35

LB 112642

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1018 WorkList ID : 145442 Department : Wet-Chemistry Date : 01-06-2021 11:17:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/05/2021	Water	M1018-01	-C pH	Cool 4 deg C	APT01	M43	WGGM-21-VWV-TSF-001	01/05/2021	SM 4500-H B
01/05/2021	Water	M1018-03	-C pH	Cool 4 deg C	APT01	M43	WGGM-21-VWV-TSF-002	01/05/2021	SM 4500-H B

Date/Time 01-06-2020 11:20
 Received by: JP Wet Chem
 Relinquished by: JP Wet Chem

Date/Time 01-06-2021 12:00
 Received by: JP Wet Chem
 Relinquished by: JP Wet Chem