



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112643

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/06/2021	09:25
2	M1008-19		1	9	100.00	>212.0	>212.0	01/06/2021	09:45
3	M1008-19DUP		1	9	100.00	>212.0	>212.0	01/06/2021	10:10
4	M1008-21		1	8	100.00	>212.0	>212.0	01/06/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112643

WorkList Name : fp1621 WorkList ID : 145426 Department : Wet-Chemistry Date : 01-06-2021 08:57:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2021	Solid	M1008-19 B	Flash Point	Cool 4 deg C	PSEG03	M12	376-384	01/05/2021	1010A
01/08/2021	Solid	M1008-21 B	Flash Point	Cool 4 deg C	PSEG03	M12	388-392	01/05/2021	1010A

Date/Time 01/06/2021 09:00
 Received by: KS (Wetchem)
 Relinquished by: AP (Ext. Lab)

Date/Time 01/06/2021 10:45
 Received by: AP C Ext - Lab 2
 Relinquished by: KS (Wetchem)