



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112644

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/06/2021	09:25
2	M1011-01		1	9	100.00	>212.0	>212.0	01/06/2021	11:15
3	M1011-01DUP		1	9	100.00	>212.0	>212.0	01/06/2021	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1621-1 WorkList ID : 145427 Department : Wet-Chemistry Date : 01-06-2021 08:58:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/12/2021	Water	M1011-01	Flash Point	Cool 4 deg C	PSEG03	M12	N48967	01/05/2021	1010A

Date/Time 01/06/2021 09:00
Received by: KS (Wetchem)
Relinquished by: AP C Ext. Lab

Date/Time 01/06/2021 11:45
Received by: AP C Ext. Lab
Relinquished by: KS (Wetchem)