



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112650

Slope : 97.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/06/2021	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/06/2021	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	01/06/2021	16:02
4	ICV	1	Water	NA	NA	20.3	7.02	01/06/2021	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/06/2021	16:06
6	M1022-01	1	Solid	20.03	20	19.8	7.01	01/06/2021	16:09
7	M1022-02	1	Solid	20.02	20	19.7	6.65	01/06/2021	16:11
8	M1022-02DUP	1	Solid	20.03	20	19.8	6.66	01/06/2021	16:12
9	CCV2	1	Water	NA	NA	20.2	12.02	01/06/2021	16:15

187 112650

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1022 WorkList ID : 145451 Department : Wet-Chemistry Date : 01-06-2021 15:28:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2021	Solid	M1022-01	pH	Cool 4 deg C	PSEG03	M21	WW-19	01/06/2021	9045D
01/06/2021	Solid	M1022-02	pH	Cool 4 deg C	PSEG03	M21	40967	01/06/2021	9045D

Date/Time 01-06-2021 15:30
 Received by: SP Log (Sim)
 Relinquished by: K-C (Sim)

Date/Time 01-06-2021 17:30
 Received by: K-C (Sim)
 Relinquished by: JD Coelt NRN