



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB112652

Analyst By : jignesh

Supervisor Review By : apatel

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/06/2021	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/06/2021	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	01/06/2021	16:02
4	ICV	1	Water	NA	NA	20.3	7.02	01/06/2021	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/06/2021	16:06
6	M1024-01	1	Water	NA	NA	13.0	6.95	01/06/2021	16:25
7	M1024-01DUP	1	Water	NA	NA	13.1	6.96	01/06/2021	16:26
8	CCV2	1	Water	NA	NA	20.2	12.02	01/06/2021	16:29

10112152

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1022 V WorkList ID : 145447 Department : Wet-Chemistry Date : 01-06-2021 15:12:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2021	Water	M1024-01	Corrosivity		PSEG03	M21	M11-BERM	01/06/2021	9040C

Date/Time 01/06/2021 15:30
 Received by: JP (SUN) (SM)
 Relinquished by: JP (SUN) (SM)

Date/Time 01-06-2021 17:13
 Received by: JC (SUN)
 Relinquished by: JP (SUN) (SM)