



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112653

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/06/2021	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/06/2021	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	01/06/2021	16:02
4	ICV	1	Water	NA	NA	20.3	7.02	01/06/2021	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/06/2021	16:06
6	M1022-05	1	Water	NA	NA	13.0	5.00	01/06/2021	16:30
7	M1022-05DUP	1	Water	NA	NA	13.1	5.01	01/06/2021	16:31
8	M1022-06	1	Water	NA	NA	12.8	4.97	01/06/2021	16:32
9	M1022-07	1	Water	NA	NA	13.6	8.20	01/06/2021	16:35
10	M1022-08	1	Water	NA	NA	13.4	4.00	01/06/2021	16:37
11	M1022-09	1	Water	NA	NA	13.6	5.82	01/06/2021	16:40
12	M1022-10	1	Water	NA	NA	13.6	7.38	01/06/2021	16:42
13	M1022-12	1	Water	NA	NA	14.0	6.13	01/06/2021	16:44
14	M1022-13	1	Water	NA	NA	14.6	6.11	01/06/2021	16:45
15	M1022-14	1	Water	NA	NA	14.5	8.58	01/06/2021	16:48
16	CCV2	1	Water	NA	NA	20.1	12.02	01/06/2021	16:50
17	M1022-15	1	Water	NA	NA	15.4	5.38	01/06/2021	16:51
18	CCV3	1	Water	NA	NA	20.1	2.01	01/06/2021	17:00

28 112653

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1022 WorkList ID : 145446 Department : Wet-Chemistry Date : 01-06-2021 15:12:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2021	Water	M1022-05	pH ☒	Cool 4 deg C	PSEG03	M21	41098	01/06/2021	9040C
01/06/2021	Water	M1022-06	pH ☒	Cool 4 deg C	PSEG03	M21	41099	01/06/2021	9040C
01/06/2021	Water	M1022-07	pH ☒	Cool 4 deg C	PSEG03	M21	41038	01/06/2021	9040C
01/06/2021	Water	M1022-08	pH ☒	Cool 4 deg C	PSEG03	M21	16790	01/06/2021	9040C
01/06/2021	Water	M1022-09	pH ☒	Cool 4 deg C	PSEG03	M21	30734	01/06/2021	9040C
01/06/2021	Water	M1022-10	pH ☒	Cool 4 deg C	PSEG03	M21	COMP-1	01/06/2021	9040C
01/06/2021	Water	M1022-12	pH ☒	Cool 4 deg C	PSEG03	M21	30732	01/06/2021	9040C
01/06/2021	Water	M1022-13	pH ☒	Cool 4 deg C	PSEG03	M21	30733	01/06/2021	9040C
01/06/2021	Water	M1022-14	pH ☒	Cool 4 deg C	PSEG03	M21	41037	01/06/2021	9040C
01/06/2021	Water	M1022-15	pH ☒	Cool 4 deg C	PSEG03	M21	41095	01/06/2021	9040C

Date/Time 01-06-2021 15:20
 Received by: J.B. (Wet Chem)
 Relinquished by: J.B. (Sun)

Date/Time 01-06-2021 17:30
 Received by: J.B. (Sun)
 Relinquished by: J.B. (Wet Chem) open