



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB112662
Analysis Date: 01/07/2021
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 01/07/2021
Extraction IN Time: 08:25
Extraction OUT Time: 09:15
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB112662BL	LB112662BL	WATER	1.3	1000	100	2.9644	2.9644	3.02	2.9644	2.9644	0.0000	0
2	LB112662BS	LB112662BS	WATER	1.3	1000	100	2.9371	2.9371	3.01	2.9557	2.9557	0.0186	18.6
3	LB112662BSD	LB112662BSD	WATER	1.3	1000	100	3.1237	3.1237	3.03	3.1446	3.1446	0.0209	20.9
4	M1011-01	N48967	WATER	1.3	1000	100	2.8999	2.8999	3.04	2.9032	2.9032	0.0033	3.3
5	M1011-01DUP	N48967DUP	WATER	1.3	1000	100	2.7411	2.7411	3.03	2.7446	2.7446	0.0035	3.5
6	M1024-01	M11-BERM	WATER	1.6	1000	100	2.8724	2.8724	3.05	2.9180	2.9180	0.0456	45.6
7	M1029-01	GRAB	WATER	1.6	1000	100	2.8543	2.8543	3.06	2.9386	2.9386	0.0843	84.3



Chemical Name	Chemical Lot #
HEXANE	W2765
PH STRIPS BOX-0.2.5	M4588
Sodium Sulfate	EP2114
1:1 HCL	WP88882
Silica Gel	W2648
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP88883
LCSWD	10.00 ML	WP88884
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-3

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>10:22</u>
1.0000 gram Balance:	<u>1.0005</u>	(0.9950-1.0050)	In Time1:	<u>09:35</u>		
Bal Check Time:	<u>08:40</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>10:45</u>
			Out Time1	<u>10:20</u>		

0.0020 gram Balance:	<u>0.0021</u> (0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>12:00</u>
1.0000 gram Balance:	<u>1.0004</u> (0.9950-1.0050)	In Time2:	<u>11:20</u>		
Bal Check Time:	<u>13:00</u>	Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>12:30</u>
		Out Time2:	<u>11:50</u>		

WORKLIST(Hardcopy Internal Chain)

LB 112662

WorkList Name : tph m1029 WorkList ID : 145459 Department : Wet-Chemistry Date : 01-07-2021 08:08:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/12/2021	Water	M1011-01	TPH	Conc H2SO4	PSEG03	M12	N48967	01/05/2021	1664A
01/12/2021	Water	M1024-01	TPH		PSEG03	M21	M11-BERM	01/06/2021	1664A
01/19/2021	Water	M1029-01	TPH	Conc H2SO4	ARAM01	M43	GRAB	01/06/2021	1664A

Date/Time 01-07-2021 8:10
 Received by: JD West Chem
 Relinquished by: CP SM1

Date/Time 01-07-2021 11:00
 Received by: CP SM1
 Relinquished by: JD West Chem