



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112665

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPONT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	01/07/2021	09:45
2	M1022-03		1	9	100.00	>212.0	>212.0	01/07/2021	10:10
3	M1022-03DUP		1	9	100.00	>212.0	>212.0	01/07/2021	10:35
4	M1022-07		1	10	100.00	>212.0	>212.0	01/07/2021	10:55
5	M1022-08		1	11	100.00	>212.0	>212.0	01/07/2021	11:25
6	M1022-09		1	12	100.00	>212.0	>212.0	01/07/2021	11:58
7	M1022-10		1	10	100.00	>212.0	>212.0	01/07/2021	12:25
8	M1024-01		1	12	100.00	>212.0	>212.0	01/07/2021	12:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112665

WorkList Name : fp1721

WorkList ID : 145471

Department : Wet-Chemistry

Date : 01-07-2021 09:12:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/13/2021	Water	M1022-03	Flash Point	Cool 4 deg C	PSEG03	M21	COMP-2	01/06/2021	1010A
01/13/2021	Water	M1022-07	Flash Point	Cool 4 deg C	PSEG03	M21	41038	01/06/2021	1010A
01/13/2021	Water	M1022-08	Flash Point	Cool 4 deg C	PSEG03	M21	16790	01/06/2021	1010A
01/13/2021	Water	M1022-09	Flash Point	Cool 4 deg C	PSEG03	M21	30734	01/06/2021	1010A
01/13/2021	Water	M1022-10	Flash Point	Cool 4 deg C	PSEG03	M21	COMP-1	01/06/2021	1010A
01/12/2021	Water	M1024-01	Flash Point		PSEG03	M21	M11-BERM	01/06/2021	1010A

Date/Time 01/07/21 09:15
 Received by: KS (Wetchem)
 Relinquished by: JPL (Wet Chem)

Date/Time 01/07/21 12:55
 Received by: JPL (Wet Chem)
 Relinquished by: KS (Wetchem)