



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 01/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB112675

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M1038-01	TC-1A	1	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1038-03	TC-1B	2	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1038-05	TC-2A	3	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1038-07	TC-2B	4	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M1038-09	CT-SP	5	1.12	9.53	7.96	81.3	
M1038-10	CT-SP	6	1.18	9.9	8.46	83.5	
M1042-01	11939	7	1.17	9.68	8.58	87.1	
M1042-02	11939	8	1.18	9.76	8.69	87.5	
M1043-01	BALL-1	9	1.11	9.82	8.76	87.8	
M1044-01	TP-2	10	1.19	9.72	9.44	96.7	
M1044-02	TP-2	11	1.12	9.57	9.3	96.8	
M1044-03	TP-4	12	1.17	9.8	9.31	94.3	
M1044-04	TP-4	13	1.16	9.92	9.41	94.2	
M1044-05	TP-6	14	1.18	9.51	9.22	96.5	
M1044-06	TP-6	15	1.17	9.57	9.32	97.0	
M1044-07	TP-8	16	1.18	9.67	9.19	94.3	
M1044-08	TP-8	17	1.19	9.64	9.26	95.5	
M1044-09	TP-10	18	1.15	9.96	9.33	92.8	
M1044-10	TP-10	19	1.17	9.71	9.1	92.9	
M1044-11	TP-11	20	1.15	9.92	9.23	92.1	
M1044-11D	TP-11DUP	21	1.19	9.94	9.25	92.1	
M1044-12	TP-11	22	1.19	9.66	8.95	91.6	
M1044-19	TP-12	23	1.14	9.98	9.11	90.2	
M1044-20	TP-12	24	1.15	9.7	8.59	87.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$