



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112682

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/08/2021	11:20
2	M1039-01		1	9	100.00	>212.0	>212.0	01/08/2021	11:45
3	M1039-01DUP		1	9	100.00	>212.0	>212.0	01/08/2021	12:15
4	M1039-03		1	10	100.00	>212.0	>212.0	01/08/2021	12:40
5	M1040-03		1	11	100.00	>212.0	>212.0	01/08/2021	13:10
6	M1040-04		1	10	100.00	>212.0	>212.0	01/08/2021	13:35
7	M1040-05		1	9	100.00	>212.0	>212.0	01/08/2021	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 112682

WorkList Name : FP 1721 WorkList ID : 145489 Department : Wet-Chemistry Date : 01-07-2021 16:12:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2021	Water	M1039-01	Flash Point	Cool 4 deg C	PSEG03	M22	41030	01/07/2021	1010A
01/14/2021	Water	M1039-03	Flash Point	Cool 4 deg C	PSEG03	M22	41031	01/07/2021	1010A
01/12/2021	Water	M1040-03	Flash Point	Cool 4 deg C	PSEG03	M23	30741	01/07/2021	1010A
01/14/2021	Water	M1040-04	Flash Point	Cool 4 deg C	PSEG03	M23	41661	01/07/2021	1010A
01/14/2021	Water	M1040-05	Flash Point	Cool 4 deg C	PSEG03	M23	41006	01/07/2021	1010A

Date/Time 01/08/21 10:55
 Received by: KS (Wetchem)
 Relinquished by: JP (Wetchem)

Date/Time 01/08/21 14:10
 Received by: JP (Wetchem)
 Relinquished by: KS (Wetchem)