


LB112690

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	49	0.000	49				
0.2	0.200	0.187	-0.013	698	0.0 %	698				
2.5	2.500	2.407	-0.093	8435	0.0 %	8435				
5.0	5.000	5.109	0.109	17859	0.0 %	17859				
7.5	7.500	7.627	0.127	26637	0.0 %	26637				
10.0	10.000	9.868	-0.132	34451	0.0 %	34451				

LB112690 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jan 2021 13:51:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	59	49	-	0.00007470A	PPB		08 Jan 2021 13:56:59	S	Std
0.2 - 1	50.2	698	-	0.20007470A	PPB		08 Jan 2021 13:59:16	S	Std
2.5 - 1	52.5	8435	-	2.50007470A	PPB		08 Jan 2021 14:01:33	S	Std
5.0 - 1	55	17859	-	5.00007470A	PPB		08 Jan 2021 14:03:50	S	Std
7.5 - 1	57.5	26637	-	7.50007470A	PPB		08 Jan 2021 14:06:07	S	Std
10.0 - 1	510	34451	-	10.00007470A	PPB		08 Jan 2021 14:08:23	S	Std
ICV66 - 1	ICV66	13715	3.9209	-7470A	PPB		08 Jan 2021 14:11:04	U	SMPL
ICB66 - 1	ICB66	-28	-0.0207	-7470A	PPB		08 Jan 2021 14:13:19	U	SMPL
CCV95 - 1	CCV95	18269	5.2270	-7470A	PPB		08 Jan 2021 14:15:34	U	SMPL
CCB95 - 1	CCB95	-23	-0.0193	-7470A	PPB		08 Jan 2021 14:17:49	U	SMPL
CRA - 1	CRA	684	0.1835	-7470A	PPB		08 Jan 2021 14:20:04	U	SMPL
HighStd - 1	HighStd	33778	9.6752	-7470A	PPB		08 Jan 2021 14:22:19	U	SMPL
ChkStd - 1	ChkStd	23935	6.8521	-7470A	PPB		08 Jan 2021 14:24:35	U	SMPL
PB134008BL - 1	PBW	-50	-0.0271	-7470A	PPB		08 Jan 2021 14:26:51	U	SMPL
PB134008BS - 1	LCSW	13184	3.7686	-7470A	PPB		08 Jan 2021 14:29:08	U	SMPL
M1019-02 - 1	20410	-16	-0.0173	-7470A	PPB		08 Jan 2021 14:31:24	U	SMPL
M1019-02DUP - 1	20410DUP	37	-0.0021	-7470A	PPB		08 Jan 2021 14:33:40	U	SMPL
M1019-02MS - 1	20410MS	13881	3.9685	-7470A	PPB		08 Jan 2021 14:35:57	U	SMPL
M1019-02MSD - 1	20410MSD	13351	3.8165	-7470A	PPB		08 Jan 2021 14:38:14	U	SMPL
M1038-11 - 1	CT-SP	-23	-0.0193	-7470A	PPB		08 Jan 2021 14:40:30	U	SMPL
CCV96 - 1	CCV96	17489	5.0033	-7470A	PPB		08 Jan 2021 14:42:45	U	SMPL
CCB96 - 1	CCB96	-39	-0.0239	-7470A	PPB		08 Jan 2021 14:45:00	U	SMPL
M1040-01 - 1	30739	34	-0.0030	-7470A	PPB		08 Jan 2021 14:47:15	U	SMPL
M1040-02 - 1	30740	27	-0.0050	-7470A	PPB		08 Jan 2021 14:49:31	U	SMPL
M1043-02 - 1	BALL-1	11	-0.0096	-7470A	PPB		08 Jan 2021 14:51:46	U	SMPL
PB134010BL - 1	PBW	19	-0.0073	-7470A	PPB		08 Jan 2021 14:54:02	U	SMPL
PB134010BS - 1	LCSW	12432	3.5529	-7470A	PPB		08 Jan 2021 14:56:17	U	SMPL
M1022-03 - 1	COMP-2	591	0.1568	-7470A	PPB		08 Jan 2021 15:03:26	U	SMPL
M1022-03DUP - 1	COMP-2DUP	542	0.1427	-7470A	PPB		08 Jan 2021 15:05:42	U	SMPL
M1022-03MS - 1	COMP-2MS	11857	3.3880	-7470A	PPB		08 Jan 2021 15:07:59	U	SMPL
M1022-03MSD - 1	COMP-2MSD	11473	3.2779	-7470A	PPB		08 Jan 2021 15:10:17	U	SMPL
M1022-07 - 1	41038	259	0.0616	-7470A	PPB		08 Jan 2021 15:12:33	U	SMPL
CCV97 - 1	CCV97	17526	5.0139	-7470A	PPB		08 Jan 2021 15:14:49	U	SMPL
CCB97 - 1	CCB97	-71	-0.0331	-7470A	PPB		08 Jan 2021 15:17:05	U	SMPL
M1022-08 - 1	16790	72	0.0079	-7470A	PPB		08 Jan 2021 15:19:20	U	SMPL
M1022-09 - 1	30734	416	0.1066	-7470A	PPB		08 Jan 2021 15:21:35	U	SMPL
M1039-01 - 1	41030	584	0.1548	-7470A	PPB		08 Jan 2021 15:23:51	U	SMPL
M1040-05 - 1	41006	365	0.0920	-7470A	PPB		08 Jan 2021 15:26:07	U	SMPL
M1058-02 - 1	PB-2021-WW-000-05	23	-0.0061	-7470A	PPB		08 Jan 2021 15:28:24	U	SMPL
PB134008TB - 1	PB133968TB	9	-0.0101	-7470A	PPB		08 Jan 2021 15:30:40	U	SMPL
M1019-02LX5 - 1		8	-0.0104	-7470A	PPB		08 Jan 2021 15:32:57	U	SMPL
M1019-02A - 1		13702	3.9172	-7470A	PPB		08 Jan 2021 15:35:14	U	SMPL
M1022-03LX5 - 1		166	0.0349	-7470A	PPB		08 Jan 2021 15:37:31	U	SMPL
M1022-03A - 1		14493	4.1440	-7470A	PPB		08 Jan 2021 15:39:47	U	SMPL
CCV98 - 1	CCV98	17575	5.0280	-7470A	PPB		08 Jan 2021 15:42:03	U	SMPL
CCB98 - 1	CCB98	-39	-0.0239	-7470A	PPB		08 Jan 2021 15:44:19	U	SMPL