



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112691

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	01/08/2021	14:35
2	CAL2	1	Water	NA	NA	20.2	7.00	01/08/2021	14:36
3	CAL3	1	Water	NA	NA	20.2	10.03	01/08/2021	14:37
4	ICV	1	Water	NA	NA	20.2	6.98	01/08/2021	14:39
5	CCV1	1	Water	NA	NA	20.2	2.01	01/08/2021	14:40
6	M1058-02	1	Water	NA	NA	17.7	10.49	01/08/2021	14:42
7	M1058-02DUP	1	Water	NA	NA	17.8	10.50	01/08/2021	14:45
8	CCV2	1	Water	NA	NA	20.2	12.02	01/08/2021	14:46

28112691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1038 V WorkList ID : 145538 Department : Wet-Chemistry Date : 01-08-2021 14:24:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/07/2021	Water	M1058-02	☒ Corrosivity	Cool 4 deg C	ARCA05	F31	PB-2021-WWW-000-05	01/07/2021	9040C

Date/Time 01-08-2021 14:30
Received by: for wet-chem
Relinquished by: KS (wet-chem)

Date/Time 01-08-2021 15:10
Received by: KS (wet-chem)
Relinquished by: for wet-chem