



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112692

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	01/08/2021	14:35
2	CAL2	1	Water	NA	NA	20.2	7.00	01/08/2021	14:36
3	CAL3	1	Water	NA	NA	20.2	10.03	01/08/2021	14:37
4	ICV	1	Water	NA	NA	20.2	6.98	01/08/2021	14:39
5	CCV1	1	Water	NA	NA	20.2	2.01	01/08/2021	14:40
6	M1058-01	1	Solid	20.03	20	20.3	7.14	01/08/2021	14:50
7	M1058-01DUP	1	Solid	20.04	20	20.2	7.15	01/08/2021	14:51
8	CCV2	1	Water	NA	NA	20.1	12.02	01/08/2021	14:53

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M1038 S WorkList ID : 145539 Department : Wet-Chemistry Date : 01-08-2021 14:25:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/07/2021	Solid	M1058-01	Corrosivity	Cool 4 deg C	ARCA05	F31	PB-2021-WS-000-05	01/07/2021	9045D

Date/Time 01-08-2021 14:25
 Received by: JS (wet Chem)
 Relinquished by: JS (wet Chem)

Date/Time 01-08-2021 15:00
 Received by: KS (wet Chem)
 Relinquished by: JS (wet Chem)