

LB 112694


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43962



A= 0.0000e+000

B= 2.2862e-004

C= 1.3624e-002

Rho= 0.9998909

Accept=Accepted

Slope
+1 on intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	e/o D
0.00	0.000	0.025	0.025	48	0.000	48					1
0.05	0.050					226					3
0.20	0.200	0.206	0.006	841	0.0 %	841					1
2.50	2.500	2.471	-0.029	10749	0.0 %	10749					1
5.00	5.000	5.038	0.038	21978	0.0 %	21978					1
7.50	7.500	7.396	-0.104	32292	0.0 %	32292					1
10.0	10.000	10.064	0.064	43962	0.0 %	43962					1

LB112694

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	48	-		0 ISM02.4	PPB	1/8/2021 16:00	Std
	0.2 S0.2	841	-		0.2 ISM02.4	PPB	1/8/2021 16:03	Std
	2.5 S2.5	10749	-		2.5 ISM02.4	PPB	1/8/2021 16:05	Std
	5 S5.0	21978	-		5 ISM02.4	PPB	1/8/2021 16:07	Std
	7.5 S7.5	32292	-		7.5 ISM02.4	PPB	1/8/2021 16:09	Std
	10 S10.0	43962	-		10 ISM02.4	PPB	1/8/2021 16:12	Std
ICV	ICV	18081	4.1473 -		ISM02.4	PPB	1/8/2021 16:15	SMPL
ICB	ICB	8	0.0155 -		ISM02.4	PPB	1/8/2021 16:17	SMPL
CCV099	CCV099	22953	5.2611 -		ISM02.4	PPB	1/8/2021 16:19	SMPL
CCB099	CCB099	-24	0.0081 -		ISM02.4	PPB	1/8/2021 16:22	SMPL
PB134007BL	PBS007	21	0.0184 -		ISM02.4	PPB	1/8/2021 16:24	SMPL
M1054-01	MBFN59	14061	3.2282 -		ISM02.4	PPB	1/8/2021 16:31	SMPL
M1054-02	MBFN59D	13593	3.1212 -		ISM02.4	PPB	1/8/2021 16:33	SMPL
M1054-03	MBFN59S	25371	5.8139 -		ISM02.4	PPB	1/8/2021 16:38	SMPL
CCV001	CCV001	22340	5.1209 -		ISM02.4	PPB	1/8/2021 16:40	SMPL
CCB001	CCB001	-23	0.0084 -		ISM02.4	PPB	1/8/2021 16:42	SMPL