



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112712

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/11/2021	11:05
2	CAL2	1	Water	NA	NA	20.1	7.00	01/11/2021	11:06
3	CAL3	1	Water	NA	NA	20.1	10.03	01/11/2021	11:07
4	ICV	1	Water	NA	NA	20.1	7.01	01/11/2021	11:11
5	CCV1	1	Water	NA	NA	20.1	2.01	01/11/2021	11:15
6	M1062-01	1	Solid	20.03	20	20.5	7.59	01/11/2021	11:22
7	M1062-01DUP	1	Solid	20.04	20	20.6	7.60	01/11/2021	11:23
8	CCV2	1	Water	NA	NA	20.2	12.02	01/11/2021	11:25

WORKLIST(Hardcopy Internal Chain)

18 112712

WorkList Name : CORROSIVITY M1063 WorkList ID : 145576 Department : Wet-Chemistry Date : 01-11-2021 10:58:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2021	Solid	M1062-01	Corrosivity	Cool 4 deg C	PSEG03	M31	HD-FDC	01/08/2021	9045D

Date/Time 01-11-2021 11:00
 Received by: JA 11062-1063
 Relinquished by: [Signature]

Date/Time 01-11-2021 12:30
 Received by: [Signature]
 Relinquished by: JO 1064 (1063)