



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112713

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/11/2021	11:05
2	CAL2	1	Water	NA	NA	20.1	7.00	01/11/2021	11:06
3	CAL3	1	Water	NA	NA	20.1	10.03	01/11/2021	11:07
4	ICV	1	Water	NA	NA	20.1	7.01	01/11/2021	11:11
5	CCV1	1	Water	NA	NA	20.1	2.01	01/11/2021	11:15
6	M1062-03	1	Water	NA	NA	15.0	6.76	01/11/2021	11:28
7	M1062-03DUP	1	Water	NA	NA	15.1	6.77	01/11/2021	11:29
8	CCV2	1	Water	NA	NA	20.2	12.02	01/11/2021	11:30

18 112713

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M1061 WorkList ID : 145577 Department : Wet-Chemistry Date : 01-11-2021 10:58:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2021	Water	M1062-03	Corrosivity	Cool 4 deg C	PSEG03	M31	HD-WD	01/08/2021	9040C

Date/Time 01-11-2021 11:20
Received by: J. A. West (JAW)
Relinquished by: J. A. West (JAW)

Date/Time 01-11-2021 12:20
Received by: J. A. West (JAW)
Relinquished by: J. A. West (JAW)