



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 01/11/2021 16:00 TEMP1 OUT: 104 °C 01/12/2021 08:00
TEMP2 IN: 103 °C 01/12/2021 08:35 TEMP2 OUT: 103 °C 01/12/2021 10:10
TEMP3 IN: 540 °C 01/12/2021 10:40 TEMP3 OUT: 540 °C 01/12/2021 12:15
TEMP4 IN: 540 °C 01/12/2021 12:40 TEMP4 OUT: 540 °C 01/12/2021 14:20

Run Number: LB112715

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB112715BL	136.8741	136.8741	100	136.8741	136.8741	136.8741	136.8741	0.0000	0
2	M1062-03	143.6041	143.6041	100	143.7271	143.7271	143.6911	143.6911	0.0360	360
3	M1062-03DUP	158.7052	158.7052	100	158.8316	158.8316	158.7958	158.7958	0.0358	358

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \times 1000000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS M1062 WorkList ID : 145579 Department : Wet-Chemistry Date : 01-11-2021 11:00:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2021	Water	M1062-03	TVS	Cool 4 deg C	PSEG03	M31	HD-WD	01/08/2021	SM2540 G

Date/Time 01-11-2021 13:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 01-11-2021 16:10
 Received by: [Signature]
 Relinquished by: [Signature]