


LB112721

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4430e-004

C= -1.9264e-002

Rho= 0.9999349

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.00	0.000	-0.010	-0.010	36	0.000	36					-
0.05	0.050					226					-10
0.20	0.200	0.181	-0.019	819	0.0 %	819					-1
2.50	2.500	2.480	-0.020	10232	0.0 %	10232					2
5.00	5.000	5.090	0.090	20914	0.0 %	20914					0
7.50	7.500	7.500	0.000	30780	0.0 %	30780					0
10.0	10.000	9.959	-0.041	40845	0.0 %	40845					0

LB112721
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	36	-		0 SFAM01.1	PPB	1/11/2021 13:48	Std
	0.2 S01	819	-		0.2 SFAM01.1	PPB	1/11/2021 13:51	Std
	2.5 S02	10232	-		2.5 SFAM01.1	PPB	1/11/2021 13:53	Std
	5 S03	20914	-		5 SFAM01.1	PPB	1/11/2021 13:55	Std
	7.5 S04	30780	-		7.5 SFAM01.1	PPB	1/11/2021 13:57	Std
	10 S05	40845	-		10 SFAM01.1	PPB	1/11/2021 14:02	Std
ICV008	ICV008	15946	3.8763 -		SFAM01.1	PPB	1/11/2021 14:05	SMPL
ICB008	ICB008	-34	-0.0276 -		SFAM01.1	PPB	1/11/2021 14:08	SMPL
CCV002	CCV002	20827	5.0687 -		SFAM01.1	PPB	1/11/2021 14:13	SMPL
CCB002	CCB002	-5	-0.0205 -		SFAM01.1	PPB	1/11/2021 14:15	SMPL
PB134035BL	PBW035	16	-0.0154 -		SFAM01.1	PPB	1/11/2021 14:17	SMPL
M1051-07	X2W75	28547	6.9547 -		SFAM01.1	PPB	1/11/2021 14:20	SMPL
M1051-08	X2W75D	28253	6.8829 -		SFAM01.1	PPB	1/11/2021 14:27	SMPL
M1051-09	X2W75S	33058	8.0567 -		SFAM01.1	PPB	1/11/2021 14:40	SMPL
CCV003	CCV003	19329	4.7028 -		SFAM01.1	PPB	1/11/2021 14:42	SMPL
CCB003	CCB003	-56	-0.0329 -		SFAM01.1	PPB	1/11/2021 14:45	SMPL