



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112731

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/11/2021	10:20
2	M1058-02		1	9	100.00	>212.0	>212.0	01/11/2021	10:45
3	M1058-02DUP		1	9	100.00	>212.0	>212.0	01/11/2021	11:10
4	M1060-01		1	11	100.00	>212.0	>212.0	01/11/2021	11:40
5	M1062-03		1	10	100.00	>212.0	>212.0	01/11/2021	12:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp11121 WorkList ID : 145560 Department : Wet-Chemistry Date : 01-11-2021 09:10:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2021	Water	M1058-02	Flash Point	Cool 4 deg C	ARCA05	F31	PB-2021-WW-000-05	01/07/2021	1010A
01/13/2021	Water	M1060-01	Flash Point	Cool 4 deg C	PSEG03	M31	N34145	01/08/2021	1010A
01/15/2021	Water	M1062-03	Flash Point	Cool 4 deg C	PSEG03	M31	HD-WD	01/08/2021	1010A

Date/Time 01/11/21 09:50
 Received by: KS (Welch)
 Relinquished by: AP C Ext. Lab)

Date/Time 01/11/21 12:15
 Received by: AP C Ext. Lab)
 Relinquished by: KS (Welch)