



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112755

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/13/2021	09:10
2	CAL2	1	Water	NA	NA	21.1	7.00	01/13/2021	09:11
3	CAL3	1	Water	NA	NA	21.1	10.03	01/13/2021	09:12
4	ICV	1	Water	NA	NA	21.1	7.00	01/13/2021	09:13
5	CCV1	1	Water	NA	NA	21.1	2.01	01/13/2021	09:15
6	M1041-02	1	Water	NA	NA	20.3	7.83	01/13/2021	09:16
7	M1041-02DUP	1	Water	NA	NA	20.4	7.84	01/13/2021	09:17
8	M1041-04	1	Water	NA	NA	21.0	7.91	01/13/2021	09:19
9	CCV2	1	Water	NA	NA	21.1	12.02	01/13/2021	09:20

✓ 112755

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1041 WorkList ID : 145645 Department : Wet-Chemistry Date : 01-13-2021 09:03:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2001	Water	M1041-02	pH	Cool 4 deg C	PSEG03	N12	50129	01/12/2021	9040C
01/29/2001	Water	M1041-04	pH	Cool 4 deg C	PSEG03	N12	30712	01/12/2021	9040C

Date/Time 01-13-2021 9:05
 Received by: JS (Wet Chem)
 Relinquished by: JS (Wet Chem)

Date/Time 01-13-2021 10:10
 Received by: JS (Wet Chem)
 Relinquished by: JS (Wet Chem)