



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112756

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/13/2021	10:38
2	CAL2	1	Water	NA	NA	21.3	7.01	01/13/2021	10:39
3	CAL3	1	Water	NA	NA	21.3	10.03	01/13/2021	10:40
4	ICV	1	Water	NA	NA	21.3	6.99	01/13/2021	10:41
5	CCV1	1	Water	NA	NA	21.3	2.01	01/13/2021	10:42
6	M1123-03	1	Solid	20.03	20	21.0	8.39	01/13/2021	10:45
7	M1123-03DUP	1	Solid	20.04	20	21.2	8.40	01/13/2021	10:46
8	CCV2	1	Water	NA	NA	21.3	12.02	01/13/2021	10:48

13112756

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1123 WorkList ID : 145658 Department : Wet-Chemistry Date : 01-13-2021 10:30:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/13/2021	Solid	M1123-03	pH	Cool 4 deg C	PSEG03	N13	VW-SP	01/13/2021	9045D

Date/Time 01-13-2021 10:35
Received by: JS (wet) (per)
Relinquished by: JS (wet)

Date/Time 01-13-2021 11:45
Received by: JS (wet)
Relinquished by: JS (wet) CR