

LB112762


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

49749



A= 0.0000e+000

B= 2.0100e-004

C= -2.8928e-003

Rho= 0.9999465

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.00	0.000	0.008	0.008	56	0.000	56					=
0.05	0.050					226					-7
0.20	0.200	0.186	-0.014	939	0.0 %	939					-1
2.50	2.500	2.469	-0.031	12299	0.0 %	12299					2
5.00	5.000	5.078	0.078	25278	0.0 %	25278					-1
7.50	7.500	7.462	-0.038	37140	0.0 %	37140					0
10.0	10.000	9.997	-0.003	49749	0.0 %	49749					

LB112762

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	56	-		0 SFAM01.1	PPB	1/13/2021 14:11	Std
	0.2 S01	939	-		0.2 SFAM01.1	PPB	1/13/2021 14:13	Std
	2.5 S02	12299	-		2.5 SFAM01.1	PPB	1/13/2021 14:16	Std
	5 S03	25278	-		5 SFAM01.1	PPB	1/13/2021 14:18	Std
	7.5 S04	37140	-		7.5 SFAM01.1	PPB	1/13/2021 14:20	Std
	10 S05	49749	-		10 SFAM01.1	PPB	1/13/2021 14:22	Std
ICV009	ICV009	18919	3.7998 -		SFAM01.1	PPB	1/13/2021 14:27	SMPL
ICB009	ICB009	10	-0.0009 -		SFAM01.1	PPB	1/13/2021 14:29	SMPL
CCV020	CCV020	25141	5.0504 -		SFAM01.1	PPB	1/13/2021 14:32	SMPL
CCB020	CCB020	-8	-0.0045 -		SFAM01.1	PPB	1/13/2021 14:34	SMPL
PB134120BL	PBS120	14	-0.0001 -		SFAM01.1	PPB	1/13/2021 14:36	SMPL
M1051-19	X2W79	88044	17.6937 -		SFAM01.1	PPB	1/13/2021 14:38	SMPL
M1051-20	X2W79D	83010	16.6819 -		SFAM01.1	PPB	1/13/2021 14:47	SMPL
M1051-21	X2W79S	99188	19.9337 -		SFAM01.1	PPB	1/13/2021 14:54	SMPL
CCV021	CCV021	25013	5.0247 -		SFAM01.1	PPB	1/13/2021 14:59	SMPL
CCB021	CCB021	-35	-0.0099 -		SFAM01.1	PPB	1/13/2021 15:01	SMPL
M1051-19DLX5	X2W79	18462	3.7079 -		SFAM01.1	PPB	1/13/2021 15:04	SMPL
M1051-20DLX5	X2W79D	17171	3.4484 -		SFAM01.1	PPB	1/13/2021 15:13	SMPL
M1051-21DLX5	X2W79S	20758	4.1694 -		SFAM01.1	PPB	1/13/2021 15:19	SMPL
CCV022	CCV022	25685	5.1597 -		SFAM01.1	PPB	1/13/2021 15:24	SMPL
CCB022	CCB022	-4	-0.0037 -		SFAM01.1	PPB	1/13/2021 15:26	SMPL