



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112770

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	01/13/2021	16:10
2	CAL2	1	Water	NA	NA	21.3	6.99	01/13/2021	16:11
3	CAL3	1	Water	NA	NA	21.3	10.03	01/13/2021	16:12
4	ICV	1	Water	NA	NA	21.3	7.00	01/13/2021	16:15
5	CCV1	1	Water	NA	NA	21.3	2.01	01/13/2021	16:16
6	M1120-01	1	Solid	20.03	20	20.4	8.73	01/13/2021	16:19
7	M1120-01DUP	1	Solid	20.02	20	20.5	8.74	01/13/2021	16:20
8	M1122-01	1	Solid	20.04	20	20.2	6.40	01/13/2021	16:22
9	CCV2	1	Water	NA	NA	21.3	12.02	01/13/2021	16:23

WORKLIST(Hardcopy Internal Chain)

VB 112770

WorkList Name : ph m1020 WorkList ID : 145672 Department : Wet-Chemistry Date : 01-13-2021 16:04:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/13/2021	Solid	M1120-01	pH	Cool 4 deg C	PSEG05	N13	AU-04-11321	01/13/2021	9045D
01/13/2021	Solid	M1122-01	pH	Cool 4 deg C	PSEG05	N22	BU-02-011321	01/13/2021	9045D

Date/Time 01-13-2021 16:05
 Received by: JP (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 01-13-2021 17:30
 Received by: KS (wetchem)
 Relinquished by: JP (wetchem)