



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112772

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/13/2021	17:15
2	CAL2	1	Water	NA	NA	21.2	7.00	01/13/2021	17:16
3	CAL3	1	Water	NA	NA	21.2	10.03	01/13/2021	17:18
4	ICV	1	Water	NA	NA	21.2	7.00	01/13/2021	17:20
5	CCV1	1	Water	NA	NA	21.3	2.01	01/13/2021	17:21
6	M1090-02	1	Solid	20.03	20	20.3	8.32	01/13/2021	17:22
7	M1090-02DUP	1	Solid	20.04	20	20.4	8.33	01/13/2021	17:23
8	M1090-05	1	Solid	20.02	20	20.3	6.45	01/13/2021	17:25
9	M1119-01	1	Solid	20.03	20	20.6	9.18	01/13/2021	17:26
10	M1121-02	1	Solid	20.04	20	20.2	7.64	01/13/2021	17:29
11	M1124-02	1	Solid	20.03	20	21.2	6.03	01/13/2021	17:30
12	M1124-04	1	Solid	20.04	20	21.0	8.60	01/13/2021	17:32
13	M1124-06	1	Solid	20.02	20	21.1	9.15	01/13/2021	17:35
14	M1125-06	1	Solid	20.01	20	20.5	8.06	01/13/2021	17:36
15	M1133-01	1	Solid	20.03	20	20.6	7.88	01/13/2021	17:39
16	CCV2	1	Water	NA	NA	21.2	12.02	01/13/2021	17:40

WORKLIST(Hardcopy Internal Chain)

112772

WorkList Name : CORROSIONITY M1119 WorkList ID : 145665 Department : Wet-Chemistry Date : 01-13-2021 11:31:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/12/2021	Solid	M1090-02	Corrosivity	Cool 4 deg C	PSEG03	N12	BORE-2-1FT-7FT	01/12/2021	9045D
01/12/2021	Solid	M1090-05	Corrosivity	Cool 4 deg C	PSEG03	N12	BORE-3-1FT-7FT	01/12/2021	9045D
01/12/2021	Solid	M1119-01	Corrosivity	Cool 4 deg C	JACO05	F61	IDW-BRI-SO-011221	01/12/2021	9045D
01/13/2021	Solid	M1121-02	Corrosivity	Cool 4 deg C	PSEG03	N21	NP-SP	01/13/2021	9045D
01/13/2021	Solid	M1124-02	Corrosivity	Cool 4 deg C	PSEG03	N21	VNJ-231	01/13/2021	9045D
01/13/2021	Solid	M1124-04	Corrosivity	Cool 4 deg C	PSEG03	N21	ETGI-358	01/13/2021	9045D
01/13/2021	Solid	M1124-06	Corrosivity	Cool 4 deg C	PSEG03	N21	VNJ-223	01/13/2021	9045D
01/13/2021	Solid	M1125-06	Corrosivity	Cool 4 deg C	PSEG03	N22	VNJ-215	01/13/2021	9045D
01/11/2021	Solid	M1133-01	Corrosivity	Cool 4 deg C	TETR16	F31	MC0AC8	01/11/2021	9045D

Date/Time 01-13-2021 16:05
 Received by: [Signature]
 Relinquished by: KS (Webster)

Date/Time 01-13-2021 18:15
 Received by: KS (Webster)
 Relinquished by: [Signature]