



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112773

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/13/2021	16:15
2	CAL2	1	Water	NA	NA	21.1	7.00	01/14/2021	16:16
3	CAL3	1	Water	NA	NA	21.1	10.03	01/13/2021	16:17
4	ICV	1	Water	NA	NA	21.2	7.01	01/13/2021	16:20
5	CCV1	1	Water	NA	NA	21.2	2.01	01/13/2021	16:21
6	M1125-07	1	Water	NA	NA	20.2	6.41	01/13/2021	16:22
7	M1125-07DUP	1	Water	NA	NA	20.3	6.42	01/13/2021	16:23
8	M1125-08	1	Water	NA	NA	20.1	11.77	01/13/2021	16:25
9	M1125-09	1	Water	NA	NA	20.1	7.40	01/13/2021	16:29
10	CCV2	1	Water	NA	NA	21.1	12.02	01/13/2021	16:30