


7471B

LB 11277 &
INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	35	0.000	35				
0.2	0.200	0.223	0.023	1071	0.0 %	1071				
2.5	2.500	2.458	-0.042	14175	0.0 %	14175				
5.0	5.000	4.947	-0.053	28764	0.0 %	28764				
7.5	7.500	7.463	-0.037	43515	0.0 %	43515				
10.0	10.000	10.063	0.063	58761	0.0 %	58761				

LB112774 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 14 Jan 2021 09:42:36					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.0000	7471B	PPB	14 Jan 2021 09:51:15	S	Std
0.2 - 1	50.2	1071	-	0.2000	7471B	PPB	14 Jan 2021 09:53:33	S	Std
2.5 - 1	52.5	14175	-	2.5000	7471B	PPB	14 Jan 2021 09:55:50	S	Std
5.0 - 1	55	28764	-	5.0000	7471B	PPB	14 Jan 2021 09:58:08	S	Std
7.5 - 1	57.5	43515	-	7.5000	7471B	PPB	14 Jan 2021 10:00:25	S	Std
10.0 - 1	510	58761	-	10.0000	7471B	PPB	14 Jan 2021 10:02:41	S	Std
ICV70 - 1	ICV70	21972	3.7881	-	7471B	PPB	14 Jan 2021 10:05:30	U	SMPL
ICB70 - 1	ICB70	-53	0.0312	-	7471B	PPB	14 Jan 2021 10:07:45	U	SMPL
CCV23 - 1	CCV23	28796	4.9521	-	7471B	PPB	14 Jan 2021 10:10:01	U	SMPL
CCB23 - 1	CCB23	-69	0.0285	-	7471B	PPB	14 Jan 2021 10:12:16	U	SMPL
CRA - 1	CRA	1070	0.2227	-	7471B	PPB	14 Jan 2021 10:14:31	U	SMPL
HighStd - 1	HighStd	58633	10.0415	-	7471B	PPB	14 Jan 2021 10:16:46	U	SMPL
ChkStd - 1	ChkStd	41629	7.1411	-	7471B	PPB	14 Jan 2021 10:19:01	U	SMPL
PB134119BL - 1	PBS	-78	0.0269	-	7471B	PPB	14 Jan 2021 10:21:19	U	SMPL
PB134119BS - 1	LCSS	20686	3.5687	-	7471B	PPB	14 Jan 2021 10:23:35	U	SMPL
M1090-02 - 1	BORE-2-1FT-7FT	1605	0.3140	-	7471B	PPB	14 Jan 2021 10:25:51	U	SMPL
M1090-02DUP - 1	BORE-2-1FT-7FTDUP	1574	0.3087	-	7471B	PPB	14 Jan 2021 10:28:08	U	SMPL
M1090-02MS - 1	BORE-2-1FT-7FTMS	22469	3.8729	-	7471B	PPB	14 Jan 2021 10:30:25	U	SMPL
M1090-02MSD - 1	BORE-2-1FT-7FTMSD	22758	3.9222	-	7471B	PPB	14 Jan 2021 10:32:42	U	SMPL
M1090-05 - 1	BORE-3-1FT-7FT	1721	0.3338	-	7471B	PPB	14 Jan 2021 10:34:57	U	SMPL
CCV24 - 1	CCV24	28567	4.9130	-	7471B	PPB	14 Jan 2021 10:37:13	U	SMPL
CCB24 - 1	CCB24	-51	0.0315	-	7471B	PPB	14 Jan 2021 10:39:28	U	SMPL
M1091-01 - 1	OK-03-011221	1236	0.2511	-	7471B	PPB	14 Jan 2021 10:41:43	U	SMPL
M1123-01 - 1	355	3095	0.5682	-	7471B	PPB	14 Jan 2021 10:43:59	U	SMPL
M1123-03 - 1	VW-SP	102951	17.6010	-	7471B	PPB	14 Jan 2021 10:46:14	U	SMPL
M1090-02LX5 - 1		87	0.0551	-	7471B	PPB	14 Jan 2021 10:48:30	U	SMPL
M1090-02A - 1		24832	4.2759	-	7471B	PPB	14 Jan 2021 10:50:48	U	SMPL
M1123-03DLX5 - 1		20321	3.5065	-	7471B	PPB	14 Jan 2021 11:03:47	U	SMPL
CCV25 - 1	CCV25	29005	4.9877	-	7471B	PPB	14 Jan 2021 11:06:03	U	SMPL
CCB25 - 1	CCB25	-34	0.0344	-	7471B	PPB	14 Jan 2021 11:08:21	U	SMPL