



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112776

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	01/14/2021	10:40
2	CAL2	1	Water	NA	NA	21.1	7.01	01/14/2021	10:41
3	CAL3	1	Water	NA	NA	21.1	10.03	01/14/2021	10:42
4	ICV	1	Water	NA	NA	21.2	6.99	01/14/2021	10:45
5	CCV1	1	Water	NA	NA	21.1	2.01	01/14/2021	10:46
6	M1144-01	1	Water	NA	NA	20.0	5.81	01/14/2021	10:48
7	M1144-01DUP	1	Water	NA	NA	20.1	5.82	01/14/2021	10:49
8	CCV2	1	Water	NA	NA	21.1	12.02	01/14/2021	10:50

103112476

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1144 WorkList ID : 145697 Department : Wet-Chemistry Date : 01-14-2021 10:33:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2021	Water	M1144-01	pH	Cool 4 deg C	STER10	F62	POOL-WATER	01/11/2021	9040C

Date/Time 01-14-2021 10:35
Received by: JOLCOY (Jem)
Relinquished by: Jem (wet-chem)

Date/Time 01-14-2021 11:35
Received by: Jem (wet-chem)
Relinquished by: Jem (wet-chem)