

Analysis Method: SM2510 BANALYST: AlexanderParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB112781Meter ID: WC CONDUCTANCE METER-1Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP86113

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB112781BL		1	19.3	0.80	0.90	01/14/2021	11:30
2	LB112781BS	1413	1	19.5	1260.00	1407.90	01/14/2021	11:33
3	M1144-01		1	19.4	5.60	6.27	01/14/2021	11:37
4	M1144-01DUP		1	19.4	5.63	6.30	01/14/2021	11:42

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

LB112781
 WORKLIST(Hardcopy Internal Chain)

WorkList Name : Conductance-11421 WorkList ID : 145695 Department : Wet-Chemistry Date : 01-14-2021 09:48:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2021	Water	M1144-01A	Conductance	Cool 4 deg C	STER10	F62	POOL-WATER	01/11/2021	SM2510 B

Date/Time 01/14/2021 11:05
 Received by: AM (wet-chem)
 Relinquished by: JD (wet-chem)

Date/Time 01/14/2021 11:55
 Received by: JD (wet-chem)
 Relinquished by: AM (wet-chem)