

LB112782


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49859



A= 0.0000e+000

B= 2.0000e-004

C= -1.6046e-002

Rho= 0.9999256

Accept=Accepted

slope
 $y = 1.9998 \times 10^{-4} x - 1.6046 \times 10^{-2}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.000	-0.000	78	0.000	78					=
0.05	0.050					226					-12
0.20	0.200	0.177	-0.023	966	0.0 %	966					-1
2.50	2.500	2.468	-0.032	12420	0.0 %	12420					2
5.00	5.000	5.093	0.093	25544	0.0 %	25544					0
7.50	7.500	7.507	0.007	37614	0.0 %	37614					0
10.0	10.000	9.956	-0.044	49859	0.0 %	49859					0

LB112782

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	78	-		0 ISM02.4	PPB	1/14/2021 12:27	Std
	0.2 S0.2	966	-		0.2 ISM02.4	PPB	1/14/2021 12:29	Std
	2.5 S2.5	12420	-		2.5 ISM02.4	PPB	1/14/2021 12:31	Std
	5 S5.0	25544	-		5 ISM02.4	PPB	1/14/2021 12:34	Std
	7.5 S7.5	37614	-		7.5 ISM02.4	PPB	1/14/2021 12:36	Std
	10 S10.0	49859	-		10 ISM02.4	PPB	1/14/2021 12:38	Std
ICV	ICV	19149	3.8138 -		ISM02.4	PPB	1/14/2021 12:41	SMPL
ICB	ICB	-5	-0.017 -		ISM02.4	PPB	1/14/2021 12:43	SMPL
CCV026	CCV026	25332	5.0504 -		ISM02.4	PPB	1/14/2021 12:46	SMPL
CCB026	CCB026	-15	-0.019 -		ISM02.4	PPB	1/14/2021 12:48	SMPL
PB134128BL	PBS128	49	-0.0062 -		ISM02.4	PPB	1/14/2021 12:50	SMPL
M1118-01	MBFN60	3630	0.71 -		ISM02.4	PPB	1/14/2021 12:52	SMPL
M1118-02	MBFN61	15880	3.16 -		ISM02.4	PPB	1/14/2021 12:55	SMPL
M1118-03	MBFN61D	17942	3.5724 -		ISM02.4	PPB	1/14/2021 12:59	SMPL
M1118-04	MBFN61S	29777	5.9394 -		ISM02.4	PPB	1/14/2021 13:04	SMPL
M1118-05	MBFN62	4713	0.9266 -		ISM02.4	PPB	1/14/2021 13:07	SMPL
CCV027	CCV027	25439	5.0718 -		ISM02.4	PPB	1/14/2021 13:09	SMPL
CCB027	CCB027	-9	-0.0178 -		ISM02.4	PPB	1/14/2021 13:11	SMPL