



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112787

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/14/2021	13:45
2	CAL2	1	Water	NA	NA	21.2	7.00	01/14/2021	13:46
3	CAL3	1	Water	NA	NA	21.2	10.03	01/14/2021	13:47
4	ICV	1	Water	NA	NA	21.2	6.99	01/14/2021	13:48
5	CCV1	1	Water	NA	NA	21.2	2.01	01/14/2021	13:49
6	M1142-02	1	Water	NA	NA	20.3	8.02	01/14/2021	13:50
7	M1142-02DUP	1	Water	NA	NA	20.4	8.03	01/14/2021	13:51
8	CCV2	1	Water	NA	NA	21.3	12.02	01/14/2021	14:00

WORKLIST(Hardcopy Internal Chain)

112727

WorkList Name : ph m1142

WorkList ID : 145706

Department : Wet-Chemistry

Date : 01-14-2021 13:39:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2021	Water	M1142-02	pH	Cool 4 deg C	PSEG03	N23	30936	01/14/2021	9040C

Date/Time 01-14-2021 13:40
 Received by: JO (wet Chem)
 Relinquished by: KS (wet Chem)

Date/Time 01-14-2021 15:00
 Received by: KS (wet Chem)
 Relinquished by: JO (wet Chem)