

LB112788

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

1/14/2021 14:28

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.984	0.0	0.111	
ICB1	0.002	0.0	0.005	
CCV1	0.981	0.0	0.111	
CCB1	-0.005	0.0	0.005	
PB134113BL	0.005	0.0	0.006	
PB134113BS	0.976	0.0	0.110	
M1101-01	-0.001	0.0	0.005	
M1101-01DUP	-0.000	0.0	0.005	
M1101-01MS	0.978	0.0	0.110	
M1101-01MSD	0.963	0.0	0.109	
CCV2	0.982	0.0	0.111	
CCB2	0.003	0.0	0.006	

N	12
Mean	0.489
SD	0.5101
CV%	104.32

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 14 13:08:26 2021

Thu Jan 14 14:26:34 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	0.0063	mg/l	1/14/2021 13:08:26	
0.05PPM	A	Phenolics-	P	0.0519	mg/l	1/14/2021 13:08:27	
0.1PPM	A	Phenolics-	P	0.1071	mg/l	1/14/2021 13:08:28	
0.25PPM	A	Phenolics-	P	0.251	mg/l	1/14/2021 13:08:29	
0.50PPM	A	Phenolics-	P	0.4952	mg/l	1/14/2021 13:08:30	
1.0PPM	A	Phenolics-	P	0.9758	mg/l	1/14/2021 13:08:31	
2.0PPM	A	Phenolics-	P	2.0128	mg/l	1/14/2021 13:08:32	
ICV1	S	Phenolics-	P	0.9836	mg/l	1/14/2021 13:43:20	
ICB1	S	Phenolics-	P	0.0016	mg/l	1/14/2021 13:43:23	
CCV1	S	Phenolics-	P	0.9814	mg/l	1/14/2021 13:43:24	
CCB1	S	Phenolics-	P	-0.0046	mg/l	1/14/2021 13:43:25	
PB134113BL	S	Phenolics-	P	0.0046	mg/l	1/14/2021 13:43:26	
PB134113BS	S	Phenolics-	P	0.9759	mg/l	1/14/2021 13:43:27	
M1101-01	S	Phenolics-	P	-0.0007	mg/l	1/14/2021 14:20:33	
M1101-01DUP	S	Phenolics-	P	-0.0003	mg/l	1/14/2021 14:20:34	
M1101-01MS	S	Phenolics-	P	0.9778	mg/l	1/14/2021 14:20:35	
M1101-01MSD	S	Phenolics-	P	0.963	mg/l	1/14/2021 14:26:30	
CCV2	S	Phenolics-	P	0.9822	mg/l	1/14/2021 14:26:32	
CCB2	S	Phenolics-	P	0.003	mg/l	1/14/2021 14:26:33	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AM

Instrument ID : Konelab

1/14/2021 13:11

Test Phenolics-


Accepted 1/14/2021 13:11

Factor 9.311

Bias 0.005

Coeff. of det. 0.999725

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors	%
1	0.00PPM	0.006	0.0063	0.0000		104
2	Phenol-2	0.011	0.0519	0.0500		107
3	Phenol-2	0.017	0.1071	0.1000		100
4	Phenol-2	0.032	0.2510	0.2500		99
5	Phenol-2	0.058	0.4952	0.5000		98
6	Phenol-2	0.110	0.9758	1.0000		101
7	Phenol-2	0.221	2.0128	2.0000		

01/14/2021