



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB112790

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	01/14/2021	14:15
2	M1041-02		1	9	100.00	>212.0	>212.0	01/14/2021	14:50
3	M1041-02DUP		1	9	100.00	>212.0	>212.0	01/14/2021	15:15
4	M1041-04		1	11	100.00	>212.0	>212.0	01/14/2021	15:45
5	M1101-01		1	10	100.00	>212.0	>212.0	01/14/2021	16:10
6	M1125-09		1	10	100.00	>212.0	>212.0	01/14/2021	16:35
7	M1142-02		1	9	100.00	>212.0	>212.0	01/14/2021	17:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB112790

WorkList Name : fp11321 WorkList ID : 145654 Department : Wet-Chemistry Date : 01-13-2021 10:16:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2001	Water	M1041-02	Flash Point	Cool 4 deg C	PSEG03	N12	50129	01/12/2021	1010A
01/29/2001	Water	M1041-04	Flash Point	Cool 4 deg C	PSEG03	N12	30712	01/12/2021	1010A
01/25/2021	Water	M1101-01	Flash Point	Cool 4 deg C	DIST01		55-ST-30	01/12/2021	1010A
01/20/2021	Water	M1125-09	Flash Point	Cool 4 deg C	PSEG03	N22	1714-1715	01/13/2021	1010A
01/28/2021	Water	M1142-02	Flash Point	Cool 4 deg C	PSEG03	N23	30936	01/14/2021	1010A

Date/Time 01/14/21 13:45
 Received by: KS (Wetchem)
 Relinquished by: RJ (EXT-196)

Date/Time 01/14/21 17:10
 Received by: RJ (EXT-196)
 Relinquished by: KS (Wetchem)