



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112795

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/14/2021	17:10
2	CAL2	1	Water	NA	NA	21.2	7.00	01/14/2021	17:11
3	CAL3	1	Water	NA	NA	21.2	10.03	01/14/2021	17:12
4	ICV	1	Water	NA	NA	21.2	7.00	01/14/2021	17:15
5	CCV1	1	Water	NA	NA	21.3	2.01	01/14/2021	17:16
6	M1146-01	1	Solid	20.03	20	20.3	7.90	01/14/2021	17:19
7	M1153-01	1	Solid	20.04	20	20.5	7.45	01/14/2021	17:20
8	M1153-01DUP	1	Solid	20.05	20	20.6	7.46	01/14/2021	17:21
9	CCV2	1	Water	NA	NA	21.2	12.02	01/14/2021	17:22

WORKLIST(Hardcopy Internal Chain)

✓ 112795

WorkList Name : PH M1153

WorkList ID : 145715

Department : Wet-Chemistry

Date : 01-14-2021 17:00:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2021	Solid	M1146-01	pH	Cool 4 deg C	ENV160	F62	520-Z	01/14/2021	9045D
01/14/2021	Solid	M1153-01	pH	Cool 4 deg C	TULL02	F32	FK-PS-01-068	01/14/2021	9045D

Date/Time 01-14-2021 17:00
 Received by: SP (watchman)
 Relinquished by: KS (watchman)

Date/Time 01-14-2021 18:00
 Received by: KS (watchman)
 Relinquished by: 20/0004