



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB112796

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/14/2021	17:10
2	CAL2	1	Water	NA	NA	21.2	7.00	01/14/2021	17:11
3	CAL3	1	Water	NA	NA	21.2	10.03	01/14/2021	17:12
4	ICV	1	Water	NA	NA	21.2	7.00	01/14/2021	17:15
5	CCV1	1	Water	NA	NA	21.3	2.01	01/14/2021	17:16
6	M1141-21	1	Solid	20.02	20	20.9	8.38	01/14/2021	17:25
7	M1141-23	1	Solid	20.01	20	20.8	10.07	01/14/2021	17:29
8	M1156-02	1	Solid	20.03	20	20.5	6.61	01/14/2021	17:32
9	M1156-02DUP	1	Solid	20.04	20	20.6	6.62	01/14/2021	17:33
10	CCV2	1	Water	NA	NA	21.4	12.02	01/14/2021	17:35

WORKLIST(Hardcopy Internal Chain)

287 112796

WorkList Name : CORROSIONITY M1141

WorkList ID : 145714

Department : Wet-Chemistry

Date : 01-14-2021 16:25:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2021	Solid	M1141-21	Corrosivity	Cool 4 deg C	PSEG03	N23	T872-T871	01/14/2021	9045D
01/14/2021	Solid	M1141-23	Corrosivity	Cool 4 deg C	PSEG03	N23	ETGI-354	01/14/2021	9045D
01/13/2021	Solid	M1156-02	Corrosivity	Cool 4 deg C	FRAN01	F42	VCP-EXC-SOIL	01/13/2021	9045D

Date/Time 01-14-2021 16:30

Received by: JO (wet Chem)

Relinquished by: KS (wetchem)

Date/Time 01-14-2021 18:00

Received by: KS (wetchem)

Relinquished by: JO (wet Chem)