


LB112804

7470A

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.4651e-004

C= -8.0998e-003

Rho= 0.9998523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.006	-0.006	9	0.000	9					-14
0.2	0.200	0.172	-0.028	731	0.0 %	731					-1
2.5	2.500	2.480	-0.020	10093	0.0 %	10093					-1
5.0	5.000	5.050	0.050	20519	0.0 %	20519					-1
7.5	7.500	7.603	0.103	30876	0.0 %	30876					-1
10.0	10.000	9.901	-0.099	40198	0.0 %	40198					-1

LB112804 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Jan 2021 09:02:58					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	7470A	PPB	15 Jan 2021 09:20:55	S	Std
0.2 - 1	50.2	731	-	0.2000	7470A	PPB	15 Jan 2021 09:23:12	S	Std
2.5 - 1	52.5	10093	-	2.5000	7470A	PPB	15 Jan 2021 09:25:30	S	Std
5.0 - 1	55	20519	-	5.0000	7470A	PPB	15 Jan 2021 09:27:47	S	Std
7.5 - 1	57.5	30876	-	7.5000	7470A	PPB	15 Jan 2021 09:30:04	S	Std
10.0 - 1	57.5	40198	-	10.0000	7470A	PPB	15 Jan 2021 09:32:20	S	Std
ICV72 - 1	ICV72	15692	3.8601	-	7470A	PPB	15 Jan 2021 09:35:05	U	SMPL
ICB72 - 1	ICB72	1	-0.0079	-	7470A	PPB	15 Jan 2021 09:37:20	U	SMPL
CCV28 - 1	CCV28	20940	5.1537	-	7470A	PPB	15 Jan 2021 09:39:35	U	SMPL
CCB28 - 1	CCB28	11	-0.0054	-	7470A	PPB	15 Jan 2021 09:41:50	U	SMPL
CRA - 1	CRA	698	0.1640	-	7470A	PPB	15 Jan 2021 09:44:05	U	SMPL
HighStd - 1	HighStd	40482	9.9709	-	7470A	PPB	15 Jan 2021 09:46:20	U	SMPL
ChkStd - 1	ChkStd	28058	6.9084	-	7470A	PPB	15 Jan 2021 09:48:36	U	SMPL
PB134152BL - 1	PBW	-102	-0.0332	-	7470A	PPB	15 Jan 2021 09:50:52	U	SMPL
PB134152BS - 1	LCSW	15495	3.8115	-	7470A	PPB	15 Jan 2021 09:53:09	U	SMPL
M1041-01 - 1	16874	293	0.0641	-	7470A	PPB	15 Jan 2021 09:55:25	U	SMPL
M1090-03 - 1	BORE-2-1FT-7FT	19	-0.0034	-	7470A	PPB	15 Jan 2021 09:57:42	U	SMPL
M1090-03DUP - 1	BORE-2-1FT-7FTDUP	72	0.0096	-	7470A	PPB	15 Jan 2021 09:59:59	U	SMPL
M1090-03MS - 1	BORE-2-1FT-7FTMS	14605	3.5921	-	7470A	PPB	15 Jan 2021 10:02:16	U	SMPL
M1090-03MSD - 1	BORE-2-1FT-7FTMSD	14641	3.6010	-	7470A	PPB	15 Jan 2021 10:04:32	U	SMPL
CCV29 - 1	CCV29	20654	5.0832	-	7470A	PPB	15 Jan 2021 10:06:47	U	SMPL
CCB29 - 1	CCB29	6	-0.0066	-	7470A	PPB	15 Jan 2021 10:09:02	U	SMPL
M1090-06 - 1	BORE-3-1FT-7FT	30	-0.0007	-	7470A	PPB	15 Jan 2021 10:11:18	U	SMPL
M1121-03 - 1	NP-SP	47	0.0035	-	7470A	PPB	15 Jan 2021 10:13:34	U	SMPL
M1123-04 - 1	VW-SP	14	-0.0046	-	7470A	PPB	15 Jan 2021 10:15:49	U	SMPL
M1124-02 - 1	VNJ-231	18	-0.0037	-	7470A	PPB	15 Jan 2021 10:18:05	U	SMPL
M1124-04 - 1	ETGI-358	83	0.0124	-	7470A	PPB	15 Jan 2021 10:20:21	U	SMPL
M1124-06 - 1	VNJ-223	71	0.0094	-	7470A	PPB	15 Jan 2021 10:22:37	U	SMPL
M1125-01 - 1	1717-1718	55	0.0055	-	7470A	PPB	15 Jan 2021 10:24:54	U	SMPL
M1125-06 - 1	VNJ-215	62	0.0072	-	7470A	PPB	15 Jan 2021 10:27:11	U	SMPL
PB134153BL - 1	PBW	29	-0.0010	-	7470A	PPB	15 Jan 2021 10:29:28	U	SMPL
PB134153BS - 1	LCSW	14300	3.5169	-	7470A	PPB	15 Jan 2021 10:31:44	U	SMPL
CCV30 - 1	CCV30	20834	5.1276	-	7470A	PPB	15 Jan 2021 10:34:00	U	SMPL
CCB30 - 1	CCB30	-3	-0.0088	-	7470A	PPB	15 Jan 2021 10:36:16	U	SMPL
M1125-07 - 1	1693-94	73	0.0099	-	7470A	PPB	15 Jan 2021 10:38:31	U	SMPL
M1125-07DUP - 1	1693-94DUP	87	0.0133	-	7470A	PPB	15 Jan 2021 10:40:47	U	SMPL
M1125-07MS - 1	1693-94MS	14665	3.6069	-	7470A	PPB	15 Jan 2021 10:43:02	U	SMPL
M1125-07MSD - 1	1693-94MSD	14986	3.6860	-	7470A	PPB	15 Jan 2021 10:50:21	U	SMPL
M1125-08 - 1	1665-1667	69	0.0089	-	7470A	PPB	15 Jan 2021 10:52:36	U	SMPL
M1125-09 - 1	1714-1715	195	0.0400	-	7470A	PPB	15 Jan 2021 10:54:52	U	SMPL
PB134152TB - 1	PB134117TB	54	0.0052	-	7470A	PPB	15 Jan 2021 10:57:09	U	SMPL
PB134153TB - 1	PB134138TB	67	0.0084	-	7470A	PPB	15 Jan 2021 10:59:26	U	SMPL
CCV31 - 1	CCV31	20822	5.1246	-	7470A	PPB	15 Jan 2021 11:01:43	U	SMPL
CCB31 - 1	CCB31	27	-0.0014	-	7470A	PPB	15 Jan 2021 11:03:59	U	SMPL
M1090-03LX5 - 1		71	0.0094	-	7470A	PPB	15 Jan 2021 11:06:15	U	SMPL
M1090-03A - 1		15770	3.8793	-	7470A	PPB	15 Jan 2021 11:08:31	U	SMPL
M1125-07LX5 - 1		36	0.0008	-	7470A	PPB	15 Jan 2021 11:10:46	U	SMPL
M1125-07A - 1		15643	3.8480	-	7470A	PPB	15 Jan 2021 11:13:02	U	SMPL
CCV32 - 1	CCV32	20612	5.0729	-	7470A	PPB	15 Jan 2021 11:15:18	U	SMPL
CCB32 - 1	CCB32	12	-0.0051	-	7470A	PPB	15 Jan 2021 11:17:33	U	SMPL