



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB112806

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2633
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	01/15/2021	11:50
2	CAL2	1	Water	NA	NA	21.2	7.00	01/15/2021	11:51
3	CAL3	1	Water	NA	NA	21.2	10.03	01/15/2021	11:52
4	ICV	1	Water	NA	NA	21.3	6.98	01/15/2021	12:00
5	CCV1	1	Water	NA	NA	21.2	2.01	01/15/2021	12:03
6	M1162-01	1	Water	NA	NA	20.1	4.82	01/15/2021	12:05
7	M1162-01DUP	1	Water	NA	NA	20.2	4.83	01/15/2021	12:06
8	M1163-01	1	Water	NA	NA	20.3	5.00	01/15/2021	12:08
9	CCV2	1	Water	NA	NA	21.3	12.02	01/15/2021	12:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M1162 WorkList ID : 145749 Department : Wet-Chemistry Date : 01-15-2021 11:43:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/11/2021	Water	M1162-01	pH	Cool 4 deg C	STER14	F61	MINI-CELL	01/11/2021	9040C
01/09/2021	Water	M1163-01	pH	Cool 4 deg C	STER11	F61	POOL-WATER	01/09/2021	9040C

Date/Time 01.15.2021 11:45
 Received by: JG/04 (Dom)
 Relinquished by: ARL (wet-chem)

Date/Time 01.15.2021 12:35
 Received by: ARL (wet-chem)
 Relinquished by: JG/04 (Dom)