

LB 112812


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49658



A= 0.0000e+000

B= 2.0048e-004

C= -5.8917e-003

Rho= 0.9999515

Accept=Accepted

slope
 $y = mx + c$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	78	0.000	78					-
0.05	0.050					226					-
0.20	0.200	0.186	-0.014	956	0.0 %	956					-7
2.50	2.500	2.470	-0.030	12350	0.0 %	12350					-1
5.00	5.000	5.025	0.025	25094	0.0 %	25094					-1
7.50	7.500	7.560	0.060	37737	0.0 %	37737					-1
10.0	10.000	9.950	-0.050	49658	0.0 %	49658					-1

LB112812

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	78	-		0 ISM02.4	PPB	1/15/2021 13:28	Std
	0.2 S0.2	956	-		0.2 ISM02.4	PPB	1/15/2021 13:30	Std
	2.5 S2.5	12350	-		2.5 ISM02.4	PPB	1/15/2021 13:32	Std
	5 S5.0	25094	-		5 ISM02.4	PPB	1/15/2021 13:34	Std
	7.5 S7.5	37737	-		7.5 ISM02.4	PPB	1/15/2021 13:37	Std
	10 S10.0	49658	-		10 ISM02.4	PPB	1/15/2021 13:39	Std
ICV	ICV	19091	3.8215 -		ISM02.4	PPB	1/15/2021 13:42	SMPL
ICB	ICB	-8	-0.0075 -		ISM02.4	PPB	1/15/2021 13:44	SMPL
CCV033	CCV033	25470	5.1004 -		ISM02.4	PPB	1/15/2021 13:46	SMPL
CCB033	CCB033	-12	-0.0083 -		ISM02.4	PPB	1/15/2021 13:48	SMPL
PB134157BL	PBS157	39	0.0019 -		ISM02.4	PPB	1/15/2021 13:51	SMPL
M1118-01	MBFN60	3610	0.7179 -		ISM02.4	PPB	1/15/2021 13:53	SMPL
M1118-02	MBFN61	17949	3.5926 -		ISM02.4	PPB	1/15/2021 13:55	SMPL
M1118-03	MBFN61D	17853	3.5733 -		ISM02.4	PPB	1/15/2021 13:59	SMPL
M1118-04	MBFN61S	30288	6.0663 -		ISM02.4	PPB	1/15/2021 14:01	SMPL
M1118-05	MBFN62	6048	1.2066 -		ISM02.4	PPB	1/15/2021 14:03	SMPL
CCV034	CCV034	25414	5.0892 -		ISM02.4	PPB	1/15/2021 14:06	SMPL
CCB034	CCB034	-9	-0.0077 -		ISM02.4	PPB	1/15/2021 14:08	SMPL