

LB 112814


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

58415



A= 0.0000e+000

B= 1.7061e-004

C= 2.7784e-002

Rho= 0.9999483

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.040	0.040	70	0.000	70					0
0.2	0.200	0.209	0.009	1062	0.0 %	1062					5
2.5	2.500	2.443	-0.057	14157	0.0 %	14157					-2
5.0	5.000	4.966	-0.034	28943	0.0 %	28943					-1
7.5	7.500	7.548	0.048	44081	0.0 %	44081					1
10.0	10.000	9.994	-0.006	58415	0.0 %	58415					0

LB112814 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 15 Jan 2021 14:49:23				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	70	-	0.00007471B	PPB	15 Jan 2021 14:50:10	S	Std
0.2 - 1	50.2	1062	-	0.20007471B	PPB	15 Jan 2021 14:52:28	S	Std
2.5 - 1	52.5	14157	-	2.50007471B	PPB	15 Jan 2021 14:54:45	S	Std
5.0 - 1	55	28943	-	5.00007471B	PPB	15 Jan 2021 14:57:02	S	Std
7.5 - 1	57.5	44081	-	7.50007471B	PPB	15 Jan 2021 14:59:19	S	Std
10.0 - 1	510	58415	-	10.00007471B	PPB	15 Jan 2021 15:01:35	S	Std
ICV73 - 1	ICV73	22522	3.8703	-7471B	PPB	15 Jan 2021 15:04:31	U	SMPL
ICB73 - 1	ICB73	-34	0.0220	-7471B	PPB	15 Jan 2021 15:06:46	U	SMPL
CCV35 - 1	CCV35	29012	4.9775	-7471B	PPB	15 Jan 2021 15:09:01	U	SMPL
CCB35 - 1	CCB35	-31	0.0225	-7471B	PPB	15 Jan 2021 15:11:17	U	SMPL
CRA - 1	CRA	1047	0.2064	-7471B	PPB	15 Jan 2021 15:13:32	U	SMPL
HighStd - 1	HighStd	59218	10.1310	-7471B	PPB	15 Jan 2021 15:15:48	U	SMPL
ChkStd - 1	ChkStd	40282	6.9003	-7471B	PPB	15 Jan 2021 15:18:03	U	SMPL
PB134147BL - 1	PBS	-89	0.0126	-7471B	PPB	15 Jan 2021 15:20:20	U	SMPL
PB134147BS - 1	LCSS	19975	3.4357	-7471B	PPB	15 Jan 2021 15:22:36	U	SMPL
M1100-02 - 1	TOPSOIL-RBF-ASHCROFT	4898	0.8634	-7471B	PPB	15 Jan 2021 15:24:52	U	SMPL
M1100-03 - 1	TOPSOIL-RBF-DARK	4511	0.7974	-7471B	PPB	15 Jan 2021 15:27:09	U	SMPL
M1100-04 - 1	TOPSOIL-RBF-LIGHT	463	0.1068	-7471B	PPB	15 Jan 2021 15:29:26	U	SMPL
M1120-01 - 1	AU-04-011321	2304	0.4209	-7471B	PPB	15 Jan 2021 15:31:43	U	SMPL
M1120-01DUP - 1	AU-04-011321DUP	1986	0.3666	-7471B	PPB	15 Jan 2021 15:33:59	U	SMPL
CCV36 - 1	CCV36	28949	4.9668	-7471B	PPB	15 Jan 2021 15:36:14	U	SMPL
CCB36 - 1	CCB36	-31	0.0225	-7471B	PPB	15 Jan 2021 15:38:29	U	SMPL
M1120-01MS - 1	AU-04-011321MS	22001	3.7814	-7471B	PPB	15 Jan 2021 15:40:45	U	SMPL
M1120-01MSD - 1	AU-04-011321MSD	22690	3.8989	-7471B	PPB	15 Jan 2021 15:43:00	U	SMPL
M1121-02 - 1	NP-SP	9120	1.5837	-7471B	PPB	15 Jan 2021 15:45:16	U	SMPL
M1122-01 - 1	BU-02-011321	877	0.1774	-7471B	PPB	15 Jan 2021 15:47:32	U	SMPL
M1124-01 - 1	VNJ-231	611	0.1320	-7471B	PPB	15 Jan 2021 15:49:48	U	SMPL
M1124-03 - 1	ETGI-358	4695	0.8288	-7471B	PPB	15 Jan 2021 15:52:04	U	SMPL
M1124-05 - 1	VNJ-223	225	0.0662	-7471B	PPB	15 Jan 2021 15:54:21	U	SMPL
M1125-05 - 1	VNJ-215	262	0.0725	-7471B	PPB	15 Jan 2021 15:56:38	U	SMPL
M1139-01 - 1	PL-01-011421	2337	0.4265	-7471B	PPB	15 Jan 2021 15:58:55	U	SMPL
M1140-01 - 1	SU-01-011421	1807	0.3361	-7471B	PPB	15 Jan 2021 16:01:11	U	SMPL
CCV37 - 1	CCV37	29038	4.9820	-7471B	PPB	15 Jan 2021 16:03:27	U	SMPL
CCB37 - 1	CCB37	-35	0.0218	-7471B	PPB	15 Jan 2021 16:05:42	U	SMPL
M1141-20 - 1	T872-T871	590	0.1284	-7471B	PPB	15 Jan 2021 16:07:58	U	SMPL
M1141-22 - 1	ETGI-354	4099	0.7271	-7471B	PPB	15 Jan 2021 16:10:14	U	SMPL
M1146-01 - 1	520-Z	2571	0.4664	-7471B	PPB	15 Jan 2021 16:12:30	U	SMPL
M1152-01 - 1	FK-EF-01-017-ABCD	5664	0.9941	-7471B	PPB	15 Jan 2021 16:14:46	U	SMPL
M1153-01 - 1	FK-PS-01-068	2832	0.5110	-7471B	PPB	15 Jan 2021 16:17:01	U	SMPL
M1156-01 - 1	VCP-EXC-SOIL	18299	3.1498	-7471B	PPB	15 Jan 2021 16:19:18	U	SMPL
M1120-01LX5 - 1		403	0.0965	-7471B	PPB	15 Jan 2021 16:21:35	U	SMPL
M1120-01A - 1		25810	4.4312	-7471B	PPB	15 Jan 2021 16:23:52	U	SMPL
CCV38 - 1	CCV38	28843	4.9487	-7471B	PPB	15 Jan 2021 16:26:09	U	SMPL
CCB38 - 1	CCB38	-42	0.0206	-7471B	PPB	15 Jan 2021 16:28:25	U	SMPL