


LB112829

7470A

INSTRUMENT ID: CV1

 μ Abs.:

40857



Linear

A= 0.0000e+000

B= 2.4905e-004

C= 2.0585e-002

Rho= 0.9992087

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.034	0.034	53	0.000	53					-
0.2	0.200	0.201	0.001	726	0.0 %	726					2
2.5	2.500	2.590	0.090	10317	0.0 %	10317					4
5.0	5.000	4.962	-0.038	19842	0.0 %	19842					-1
7.5	7.500	7.217	-0.283	28895	0.0 %	28895					-4
10.0	10.000	10.196	0.196	40857	0.0 %	40857					2

LB112829 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jan 2021 11:52:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	53	-	0.0000	7470A	PPB	18 Jan 2021 11:57:16	S	Std
0.2 - 1	50.2	726	-	0.2000	7470A	PPB	18 Jan 2021 11:59:33	S	Std
2.5 - 1	52.5	10317	-	2.5000	7470A	PPB	18 Jan 2021 12:01:50	S	Std
5.0 - 1	55	19842	-	5.0000	7470A	PPB	18 Jan 2021 12:04:07	S	Std
7.5 - 1	57.5	28895	-	7.5000	7470A	PPB	18 Jan 2021 12:06:23	S	Std
10.0 - 1	51.8	40857	-	10.0000	7470A	PPB	18 Jan 2021 12:13:46	S	Std
ICV74 - 1	ICV74	14517	3.6360	-	7470A	PPB	18 Jan 2021 12:16:23	U	SMPL
ICB74 - 1	ICB74	-39	0.0109	-	7470A	PPB	18 Jan 2021 12:18:38	U	SMPL
CCV39 - 1	CCV39	20242	5.0618	-	7470A	PPB	18 Jan 2021 12:20:53	U	SMPL
CCB39 - 1	CCB39	-22	0.0151	-	7470A	PPB	18 Jan 2021 12:23:08	U	SMPL
CRA - 1	CRA	698	0.1944	-	7470A	PPB	18 Jan 2021 12:25:23	U	SMPL
HighStd - 1	HighStd	40847	10.1934	-	7470A	PPB	18 Jan 2021 12:27:39	U	SMPL
ChkStd - 1	ChkStd	27969	6.9862	-	7470A	PPB	18 Jan 2021 12:29:54	U	SMPL
PB134184BL - 1	PBW	-58	0.0061	-	7470A	PPB	18 Jan 2021 12:32:10	U	SMPL
PB134184BS - 1	LCSW	15875	3.9742	-	7470A	PPB	18 Jan 2021 12:34:26	U	SMPL
M1141-21 - 1	T872-T871	-43	0.0099	-	7470A	PPB	18 Jan 2021 12:36:43	U	SMPL
M1141-21DUP - 1	T872-T871DUP	25	0.0268	-	7470A	PPB	18 Jan 2021 12:38:59	U	SMPL
M1141-21MS - 1	T872-T871MS	14153	3.5453	-	7470A	PPB	18 Jan 2021 12:41:16	U	SMPL
M1141-21MSD - 1	T872-T871MSD	14210	3.5595	-	7470A	PPB	18 Jan 2021 12:43:32	U	SMPL
M1141-23 - 1	ETGI-354	-24	0.0146	-	7470A	PPB	18 Jan 2021 12:45:48	U	SMPL
CCV40 - 1	CCV40	19996	5.0005	-	7470A	PPB	18 Jan 2021 12:48:03	U	SMPL
CCB40 - 1	CCB40	-14	0.0171	-	7470A	PPB	18 Jan 2021 12:50:19	U	SMPL
M1146-02 - 1	520-Z	45	0.0318	-	7470A	PPB	18 Jan 2021 12:52:34	U	SMPL
M1156-02 - 1	VCP-EXC-SOIL	38	0.0300	-	7470A	PPB	18 Jan 2021 12:54:50	U	SMPL
PB134184TB - 1	PB134158TB	30	0.0281	-	7470A	PPB	18 Jan 2021 12:57:05	U	SMPL
M1141-21LX5 - 1		36	0.0296	-	7470A	PPB	18 Jan 2021 12:59:21	U	SMPL
M1141-21A - 1		17100	4.2793	-	7470A	PPB	18 Jan 2021 13:01:37	U	SMPL
CCV41 - 1	CCV41	20242	5.0618	-	7470A	PPB	18 Jan 2021 13:03:54	U	SMPL
CCB41 - 1	CCB41	-18	0.0161	-	7470A	PPB	18 Jan 2021 13:06:10	U	SMPL