



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112840

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP89164
Turbidity Calibration std, 20NTU	WP89153
Turbidity Calibration std, 40NTU	WP89152
Turbidity Calibration std, 1NTU	WP89166
Turbidity Calibration std, 5NTU	WP89163
Turbidity Calibration - CCV std, 10 NTU	WP89170
10 NTU Standard 500 ml	W2752

Intercept: 0.1816

Slope: 0.977

Regression: 0.9997

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.245	01/18/2021	15:40
2	CAL2	1	1	1.150	01/18/2021	15:42
3	CAL3	5	1	5.130	01/18/2021	15:44
4	CAL4	10	1	9.420	01/18/2021	15:46
5	CAL5	20	1	20.300	01/18/2021	15:48
6	CAL6	40	1	39.100	01/18/2021	15:51



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB112840

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.500	10.561	01/18/2021	15:53
2	ICB		1	0.276	0.097	01/18/2021	15:55
3	CCV1	10	1	9.820	9.865	01/18/2021	15:57
4	CCB1		1	0.287	0.108	01/18/2021	16:00
5	LB112840BL		1	0.294	0.115	01/18/2021	16:02
6	M1126-01		1	6.870	6.846	01/18/2021	16:05
7	CCV2	10	1	9.920	9.968	01/18/2021	16:07
8	CCB2		1	0.281	0.102	01/18/2021	16:10