

7470A


LB112844

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43752



A= 0.0000e+000

B= 2.3269e-004

C= -6.0817e-002

Rho= 0.9993902

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	-0.024	-0.024	157	0.000	157					11
0.2	0.200	0.221	0.021	1213	0.0 %	1213					1
2.5	2.500	2.525	0.025	11111	0.0 %	11111					2
5.0	5.000	5.121	0.121	22268	0.0 %	22268					-4
7.5	7.500	7.237	-0.263	31364	0.0 %	31364					1
10.0	10.000	10.120	0.120	43752	0.0 %	43752					

LB112844 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Jan 2021 11:09:26						
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type	
0.0 - 1	50	157	-	0.0000	7470A	PPB	19 Jan 2021 11:16:46	S	Std	
0.2 - 1	50.2	1213	-	0.2000	7470A	PPB	19 Jan 2021 11:21:43	S	Std	
2.5 - 1	52.5	11111	-	2.5000	7470A	PPB	19 Jan 2021 11:24:00	S	Std	
5.0 - 1	55	22268	-	5.0000	7470A	PPB	19 Jan 2021 11:26:17	S	Std	
7.5 - 1	57.5	31364	-	7.5000	7470A	PPB	19 Jan 2021 11:28:34	S	Std	
10.0 - 1	50	43752	-	10.0000	7470A	PPB	19 Jan 2021 11:35:52	S	Std	
ICV76 - 1	ICV76	16997	3.8943		7470A	PPB	19 Jan 2021 11:38:31	U	SMPL	
ICB76 - 1	ICB76	295	0.0078		7470A	PPB	19 Jan 2021 11:40:46	U	SMPL	
CCV46 - 1	CCV46	22722	5.2265		7470A	PPB	19 Jan 2021 11:43:01	U	SMPL	
CCB46 - 1	CCB46	276	0.0034		7470A	PPB	19 Jan 2021 11:45:16	U	SMPL	
CRA - 1	CRA	1201	0.2186		7470A	PPB	19 Jan 2021 11:47:32	U	SMPL	
HighStd - 1	HighStd	43406	10.0395		7470A	PPB	19 Jan 2021 11:49:47	U	SMPL	
ChkStd - 1	ChkStd	30491	7.0342		7470A	PPB	19 Jan 2021 11:52:02	U	SMPL	
PB134185BL - 1	PBW	170	-0.0213		7470A	PPB	19 Jan 2021 11:57:51	U	SMPL	
PB134185BS - 1	LCSW	15436	3.5310		7470A	PPB	19 Jan 2021 12:02:44	U	SMPL	
M1082-04 - 1	BACKFILL-ASHCROFT-01	1259	0.2321		7470A	PPB	19 Jan 2021 12:05:00	U	SMPL	
M1082-05 - 1	BACKFILL-ASHCROFT-02	883	0.1447		7470A	PPB	19 Jan 2021 12:07:16	U	SMPL	
M1100-05 - 1	TOPSOIL-RBF-ASHCROFT	68	-0.0450		7470A	PPB	19 Jan 2021 12:09:34	U	SMPL	
M1100-05DUP - 1	TOPSOIL-RBF-ASHCROFTDUP	87	-0.0406		7470A	PPB	19 Jan 2021 12:11:51	U	SMPL	
M1100-05MS - 1	TOPSOIL-RBF-ASHCROFTMS	15462	3.5371		7470A	PPB	19 Jan 2021 12:14:07	U	SMPL	
CCV47 - 1	CCV47	22308	5.1301		7470A	PPB	19 Jan 2021 12:16:22	U	SMPL	
CCB47 - 1	CCB47	89	-0.0401		7470A	PPB	19 Jan 2021 12:18:38	U	SMPL	
M1100-05MSD - 1	TOPSOIL-RBF-ASHCROFTMSD	14418	3.2942		7470A	PPB	19 Jan 2021 12:20:53	U	SMPL	
M1100-06 - 1	TOPSOIL-RBF-DARK	43	-0.0508		7470A	PPB	19 Jan 2021 12:23:08	U	SMPL	
M1100-07 - 1	TOPSOIL-RBF-LIGHT	99	-0.0378		7470A	PPB	19 Jan 2021 12:25:24	U	SMPL	
PB134202BL - 1	PBW	290	0.0067		7470A	PPB	19 Jan 2021 12:27:40	U	SMPL	
PB134202BS - 1	LCSW	15912	3.6418		7470A	PPB	19 Jan 2021 12:29:56	U	SMPL	
M1076-05 - 1	OR-03-011121	673	0.0958		7470A	PPB	19 Jan 2021 12:32:12	U	SMPL	
M1159-05 - 1	BORE-2-ALT	886	0.1453		7470A	PPB	19 Jan 2021 12:34:28	U	SMPL	
M1159-05DUP - 1	BORE-2-ALTDUP	1001	0.1721		7470A	PPB	19 Jan 2021 12:36:46	U	SMPL	
M1159-05MS - 1	BORE-2-ALTMS	16693	3.8235		7470A	PPB	19 Jan 2021 12:39:02	U	SMPL	
M1159-05MSD - 1	BORE-2-ALTMSD	16948	3.8829		7470A	PPB	19 Jan 2021 12:41:18	U	SMPL	
CCV48 - 1	CCV48	22244	5.1152		7470A	PPB	19 Jan 2021 12:43:34	U	SMPL	
CCB48 - 1	CCB48	152	-0.0254		7470A	PPB	19 Jan 2021 12:45:49	U	SMPL	
PB134203BL - 1	PBW	105	-0.0364		7470A	PPB	19 Jan 2021 12:53:28	U	SMPL	
PB134203BS - 1	LCSW	17346	3.9755		7470A	PPB	19 Jan 2021 12:55:44	U	SMPL	
M1142-02 - 1	30936	920	0.1533		7470A	PPB	19 Jan 2021 12:57:59	U	SMPL	
M1142-02DUP - 1	30936DUP	932	0.1561		7470A	PPB	19 Jan 2021 13:00:15	U	SMPL	
M1142-02MS - 1	30936MS	14898	3.4059		7470A	PPB	19 Jan 2021 13:02:31	U	SMPL	
M1142-02MSD - 1	30936MSD	14869	3.3991		7470A	PPB	19 Jan 2021 13:04:47	U	SMPL	
PB134185TB - 1	PB134151TB	430	0.0392		7470A	PPB	19 Jan 2021 13:09:44	U	SMPL	
PB134202TB - 1	PB134188TB	637	0.0874		7470A	PPB	19 Jan 2021 13:12:01	U	SMPL	
PB134203TB - 1	PB134189TB	648	0.0900		7470A	PPB	19 Jan 2021 13:17:49	U	SMPL	
CCV49 - 1	CCV49	22424	5.1571		7470A	PPB	19 Jan 2021 13:20:04	U	SMPL	
CCB49 - 1	CCB49	93	-0.0392		7470A	PPB	19 Jan 2021 13:22:20	U	SMPL	
M1100-05LX5 - 1		0	-0.0608		7470A	PPB	19 Jan 2021 13:24:36	U	SMPL	
M1100-05A - 1		16720	3.8298		7470A	PPB	19 Jan 2021 13:26:51	U	SMPL	
M1159-05LX5 - 1		99	-0.0378		7470A	PPB	19 Jan 2021 13:29:07	U	SMPL	
M1159-05A - 1		16628	3.8084		7470A	PPB	19 Jan 2021 13:31:23	U	SMPL	
M1142-02LX5 - 1		102	-0.0371		7470A	PPB	19 Jan 2021 13:33:39	U	SMPL	
M1142-02A - 1		18463	4.2354		7470A	PPB	19 Jan 2021 13:35:54	U	SMPL	
CCV50 - 1	CCV50	21039	4.8348		7470A	PPB	19 Jan 2021 13:38:11	U	SMPL	
CCB50 - 1	CCB50	73	-0.0438		7470A	PPB	19 Jan 2021 13:40:27	U	SMPL	