

LB 112848


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45927



A= 0.0000e+000

B= 2.1708e-004

C= -1.7038e-002

Rho= 0.9998019

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	255	0.000	255					17
0.05	0.050					226					6
0.20	0.200	0.234	0.034	1156	0.0 %	1156					1
2.50	2.500	2.360	-0.140	10951	0.0 %	10951					1
5.00	5.000	5.031	0.031	23256	0.0 %	23256					1
7.50	7.500	7.584	0.084	35013	0.0 %	35013					1
10.0	10.000	9.953	-0.047	45927	0.0 %	45927					0

LB112848
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	255	-		0 ISM02.4	PPB	1/19/2021 14:31	Std
	0.2 S0.2	1156	-		0.2 ISM02.4	PPB	1/19/2021 14:33	Std
	2.5 S2.5	10951	-		2.5 ISM02.4	PPB	1/19/2021 14:36	Std
	5 S5.0	23256	-		5 ISM02.4	PPB	1/19/2021 14:40	Std
	7.5 S7.5	35013	-		7.5 ISM02.4	PPB	1/19/2021 14:43	Std
	10 S10.0	45927	-		10 ISM02.4	PPB	1/19/2021 14:45	Std
ICV	ICV	17483	3.7781 -		ISM02.4	PPB	1/19/2021 14:48	SMPL
ICB	ICB	186	0.0233 -		ISM02.4	PPB	1/19/2021 14:50	SMPL
CCV051	CCV051	23555	5.0962 -		ISM02.4	PPB	1/19/2021 14:52	SMPL
CCB051	CCB051	168	0.0194 -		ISM02.4	PPB	1/19/2021 14:55	SMPL
PB134221BL	PBW221	34	-0.0097 -		ISM02.4	PPB	1/19/2021 14:59	SMPL
M1130-01	MA42Y8	471	0.0852 -		ISM02.4	PPB	1/19/2021 15:06	SMPL
M1130-02	MA42Y8D	439	0.0783 -		ISM02.4	PPB	1/19/2021 15:09	SMPL
M1130-03	MA42Y8S	4810	1.0271 -		ISM02.4	PPB	1/19/2021 15:11	SMPL
PB134222BL	PBW222	49	-0.0064 -		ISM02.4	PPB	1/19/2021 15:13	SMPL
M1145-08	MF1A17	230	0.0329 -		ISM02.4	PPB	1/19/2021 15:15	SMPL
M1145-20	MF1A16	232	0.0333 -		ISM02.4	PPB	1/19/2021 15:18	SMPL
CCV052	CCV052	23138	5.0057 -		ISM02.4	PPB	1/19/2021 15:20	SMPL
CCB052	CCB052	42	-0.0079 -		ISM02.4	PPB	1/19/2021 15:22	SMPL