

LB112885

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AM

Instrument ID : Konelab

1/20/2021 11:05

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.997	0.0	0.114	
ICB1	0.000	0.0	0.006	
CCV1	0.983	0.0	0.112	
CCB1	-0.001	0.0	0.006	
PB134229BL	-0.003	0.0	0.005	
PB134229BS	0.974	0.0	0.112	
M1175-01	0.165	0.0	0.024	
M1175-01DUP	0.173	0.0	0.024	
M1175-01MS	1.130	0.0	0.129	
M1175-01MSD	1.136	0.0	0.129	
CCV2	0.981	0.0	0.112	
CCB2	-0.004	0.0	0.005	

N	12
Mean	0.544
SD	0.5172
CV%	95.01

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 20 09:38:43 2021

Wed Jan 20 11:02:55 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	-0.0013	mg/l	1/20/2021 9:38:43	
0.05PPM	A	Phenolics-	P	0.0535	mg/l	1/20/2021 9:38:44	
0.1PPM	A	Phenolics-	P	0.1078	mg/l	1/20/2021 9:38:45	
0.25PPM	A	Phenolics-	P	0.2525	mg/l	1/20/2021 9:38:46	
0.50PPM	A	Phenolics-	P	0.4997	mg/l	1/20/2021 9:38:47	
1.0PPM	A	Phenolics-	P	0.9771	mg/l	1/20/2021 9:38:48	
2.0PPM	A	Phenolics-	P	2.0107	mg/l	1/20/2021 9:38:49	
ICV1	S	Phenolics-	P	0.9974	mg/l	1/20/2021 10:09:05	
ICB1	S	Phenolics-	P	0.0001	mg/l	1/20/2021 10:09:07	
CCV1	S	Phenolics-	P	0.9829	mg/l	1/20/2021 10:09:09	
CCB1	S	Phenolics-	P	-0.001	mg/l	1/20/2021 10:09:10	
PB134229BL	S	Phenolics-	P	-0.0026	mg/l	1/20/2021 10:09:12	
PB134229BS	S	Phenolics-	P	0.9745	mg/l	1/20/2021 10:09:13	
M1175-01	S	Phenolics-	P	0.1653	mg/l	1/20/2021 11:02:45	
M1175-01DUP	S	Phenolics-	P	0.1726	mg/l	1/20/2021 11:02:47	
M1175-01MS	S	Phenolics-	P	1.1304	mg/l	1/20/2021 11:02:49	
M1175-01MSD	S	Phenolics-	P	1.1355	mg/l	1/20/2021 11:02:50	
CCV2	S	Phenolics-	P	0.9813	mg/l	1/20/2021 11:02:53	
CCB2	S	Phenolics-	P	-0.0042	mg/l	1/20/2021 11:02:54	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AM

Instrument ID : Konelab

1/20/2021 9:41


Test Phenolics-

Accepted 1/20/2021 9:41

Factor 9.202
Bias 0.006

Coeff. of det. 0.999772

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors	%
1	0.00PPM	0.006	-0.0013	0.0000		107
2	Phenol-2	0.012	0.0535	0.0500		108
3	Phenol-2	0.017	0.1078	0.1000		101
4	Phenol-2	0.033	0.2525	0.2500		100
5	Phenol-2	0.060	0.4997	0.5000		98
6	Phenol-2	0.112	0.9771	1.0000		101
7	Phenol-2	0.224	2.0107	2.0000		

01/20/2021
AM