


LB112858

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9850e-004

C= -3.2863e-002

Rho= 0.9998784

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.007	0.007	199	0.000	199					-
0.05	0.050					226					-
0.20	0.200	0.177	-0.023	1059	0.0 %	1059					-12
2.50	2.500	2.472	-0.028	12618	0.0 %	12618					-1
5.00	5.000	5.026	0.026	25484	0.0 %	25484					-1
7.50	7.500	7.603	0.103	38469	0.0 %	38469					-1
10.0	10.000	9.915	-0.085	50116	0.0 %	50116					-1

LB112858
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	199	-		0 ISM02.4	PPB	1/20/2021 13:37	Std
	0.2 S0.2	1059	-		0.2 ISM02.4	PPB	1/20/2021 13:40	Std
	2.5 S2.5	12618	-		2.5 ISM02.4	PPB	1/20/2021 13:42	Std
	5 S5.0	25484	-		5 ISM02.4	PPB	1/20/2021 13:44	Std
	7.5 S7.5	38469	-		7.5 ISM02.4	PPB	1/20/2021 13:47	Std
	10 S10.0	50116	-		10 ISM02.4	PPB	1/20/2021 13:49	Std
ICV	ICV	19594	3.8566 -		ISM02.4	PPB	1/20/2021 13:55	SMPL
ICB	ICB	138	-0.0055 -		ISM02.4	PPB	1/20/2021 13:57	SMPL
CCV053	CCV053	25636	5.0559 -		ISM02.4	PPB	1/20/2021 13:59	SMPL
CCB053	CCB053	132	-0.0067 -		ISM02.4	PPB	1/20/2021 14:01	SMPL
PB134240BL	PBW240	52	-0.0225 -		ISM02.4	PPB	1/20/2021 14:04	SMPL
M1145-01	MF1A08	1848	0.334 -		ISM02.4	PPB	1/20/2021 14:06	SMPL
M1145-02	MF1A09	274	0.0215 -		ISM02.4	PPB	1/20/2021 14:08	SMPL
M1145-03	MF1A10	930	0.1517 -		ISM02.4	PPB	1/20/2021 14:10	SMPL
M1145-04	MF1A11	230	0.0128 -		ISM02.4	PPB	1/20/2021 14:13	SMPL
M1145-05	MF1A12	223	0.0114 -		ISM02.4	PPB	1/20/2021 14:15	SMPL
M1145-06	MF1A13	316	0.0299 -		ISM02.4	PPB	1/20/2021 14:17	SMPL
M1145-07	MF1A15	1161	0.1976 -		ISM02.4	PPB	1/20/2021 14:20	SMPL
M1145-09	MF1A00	20790	4.094 -		ISM02.4	PPB	1/20/2021 14:22	SMPL
M1145-10	MF1A01	238	0.0144 -		ISM02.4	PPB	1/20/2021 14:24	SMPL
M1145-11	MF1A01D	203	0.0074 -		ISM02.4	PPB	1/20/2021 14:26	SMPL
M1145-12	MF1A01S	12460	2.4405 -		ISM02.4	PPB	1/20/2021 14:29	SMPL
M1145-13	MF1A02	735	0.113 -		ISM02.4	PPB	1/20/2021 14:31	SMPL
M1145-14	MF1A03	391	0.0448 -		ISM02.4	PPB	1/20/2021 14:33	SMPL
M1145-15	MF1A04	3080	0.5785 -		ISM02.4	PPB	1/20/2021 14:35	SMPL
M1145-16	MF1A05	352	0.037 -		ISM02.4	PPB	1/20/2021 14:38	SMPL
M1145-17	MF1A06	3688	0.6992 -		ISM02.4	PPB	1/20/2021 14:40	SMPL
M1145-18	MF1A07	284	0.0235 -		ISM02.4	PPB	1/20/2021 14:42	SMPL
M1145-19	MF1A14	783	0.1226 -		ISM02.4	PPB	1/20/2021 14:44	SMPL
CCV054	CCV054	25600	5.0488 -		ISM02.4	PPB	1/20/2021 14:47	SMPL
CCB054	CCB054	101	-0.0128 -		ISM02.4	PPB	1/20/2021 14:49	SMPL